

Analyzing the Effects of Capacitances-to-Shield in Sample Probes on AC Quantized Hall Resistance Measurements

Volume 104

Number 4

July–August 1999

M. E. Cage and A. Jeffery

National Institute of Standards and Technology,
Gaithersburg, MD 20899-0001

We analyze the effects of the large capacitances-to-shields existing in all sample probes on measurements of the ac quantized Hall resistance R_H . The object of this analysis is to investigate how these capacitances affect the observed frequency dependence of R_H . Our goal is to see if there is some way to eliminate or minimize this significant frequency dependence, and thereby realize an intrinsic ac quantized Hall resistance standard. Equivalent electrical circuits are used in this analysis, with circuit components consisting of: capacitances and leakage resistances to the sample probe shields; inductances and resistances of the sample probe leads; quantized Hall resistances, longitudinal resistances, and voltage generators within the quantum Hall effect device; and multiple connections to the device. We derive exact algebraic equations for the measured

R_H values expressed in terms of the circuit components. Only two circuits (with single-series “offset” and quadruple-series connections) appear to meet our desired goals of measuring both R_H and the longitudinal resistance R_x in the same cool-down for both ac and dc currents with a one-standard-deviation uncertainty of $10^{-8} R_H$ or less. These two circuits will be further considered in a future paper in which the effects of wire-to-wire capacitances are also included in the analysis.

Key words: ac quantum Hall effect; capacitance-to-shield; equivalent electrical circuit; longitudinal resistance; multi-series connections; quantized Hall resistance.

Accepted: June 2, 1999

Available online: <http://www.nist.gov/jres>

1. Introduction

Many laboratories are now attempting to employ the integer quantum Hall effect (QHE) [1–3] to realize an intrinsic ac resistance standard by using ac bridges to compare ac quantized Hall resistances R_H with ac reference standards. In experiments reported to date [4–9], the measured values of the ac quantized Hall resistances R_H unfortunately varied with the applied frequency f of the current, and differed from the dc value of R_H by at least $10^{-7} R_H$ at a frequency f of 1592 Hz (where the angular frequency $\omega = 2\pi f$ is 10^4 rad/s). Furthermore, some sample probe leads had to be removed at the device in order to reduce the frequency dependence to

this still significant amount. Lead removal creates two problems: (1) parasitic impedances within the QHE resistance standard (which arise from capacitances, inductances, lead resistances, and leakage resistances) become more difficult to measure or estimate, making it harder to apply corrections to the measured values of R_H ; and (2) measurements of both R_H and the longitudinal resistance R_x can not be made during the same cool-down, which has been found to be necessary [10] in order to obtain reliable values of R_H with direct (dc) currents.

Our desired goal at NIST is to measure both R_H and the longitudinal resistance R_x in the same cool-down for both ac and dc currents with all sample probe leads attached, and to do this with a one-standard-deviation uncertainty equal to or less than $10^{-8} R_H$ in order to verify and replace parts of the calculable capacitor chain [11] that provides the System International (SI) value of R_H at NIST. The one-standard-deviation uncertainty of the entire NIST calculable capacitor chain is $2.4 \times 10^{-8} R_H$. Therefore, we need to achieve uncertainties of $10^{-8} R_H$ or less in the ac R_H measurements.

Therefore, the frequency dependence of R_H is a serious problem that must be addressed. This paper investigates the effects of the capacitances-to-shield, and the series inductances and series resistances of sample probe leads on measurements of the ac R_H . It also identifies ways to eliminate or minimize the frequency dependences resulting from these parasitic impedances. Most of the capacitances-to-shield arise from the capacitances between the inner and outer conductors of the coaxial leads and connectors within the ac quantized Hall resistance standard; a smaller amount arises from the capacitances between the quantum Hall effect device plus sample holder and the surrounding conducting surfaces of the sample probe.

2. Strategy

We investigate the effects of capacitances-to-shield on measurements of R_H by using equivalent electrical circuits and multiple connections to the quantum Hall effect device. The multiple connections will be defined in Secs. 7–9. We derive exact algebraic equations for the currents and quantum Hall voltages of the standard. The discrete circuit components consist of: (a) capacitances and leakage resistances to the shields of the ac quantized Hall resistance standard; (b) inductances and series resistances of the internal and external sample probe leads and connectors; and (c) quantized Hall resistances, longitudinal resistances, and voltage generators within the quantum Hall effect device itself. These circuit components include everything within the standard except wire-to-wire capacitances between pairs of the inner conductors. Significant wire-to-wire capacitances can exist between pairs of conducting surfaces of the quantum Hall effect device, the sample holder, and the bonding wires between them. The wire-to-wire capacitances may be important, but their inclusion makes the circuit analyses extremely difficult, so they are excluded at this intermediate stage where we are trying to find viable circuit candidates for the final analysis of a complete equivalent circuit representation of an ac quantized Hall resistance standard.

We give a brief explanation of the dc quantum Hall effect in Sec. 3. Section 4 describes our equivalent electrical circuit model of an ac quantized Hall resistance standard. Single-series “normal”, single-series “offset”, double-series, and quadruple-series circuits are explained and analyzed in Secs. 5–7 and Sec. 9. We find that two of these circuits (those with single-series “offset” and quadruple-series connections) appear to meet our desired goals of measuring both R_H and the longitudinal resistance R_x in the same cool-down for both ac and dc currents with an uncertainty of $10^{-8} R_H$ or less. These two circuits will be analyzed in more detail in a future paper in which the effects of wire-to-wire capacitances are also included in the analysis.

3. DC Quantum Hall Effect

The quantum Hall effect (QHE) has been successfully used as an intrinsic dc resistance standard. In the integer dc QHE [1–3], the Hall resistance R_H of the i th plateau of a fully-quantized, two-dimensional electron gas (2DEG) is $R_H(i) = V_H(i)/I_T$, where $V_H(i)$ is the quantum Hall voltage measured between potential probes located on opposite sides of the device, and I_T is the total current flowing between the source and drain current contacts at the ends of the device. Under ideal conditions, the values of $R_H(i)$ in standards-quality devices satisfy the relationships $R_H(i) = h/(e^2 i) = R_K/i$, where h is the Planck constant, e is the elementary charge, i is an integer, and R_K is the von Klitzing constant, $R_K \approx 25\,812.807\, \Omega$ [12]. However, the conditions are not always ideal. The values of $R_H(i)$ can vary with the device temperature T [13] and with the applied current I_T [14]. Thus the measured dc values of $R_H(i)$ are not necessarily equal to $h/(e^2 i)$.

The current flow within the 2DEG is nearly dissipationless in the quantum Hall plateau regions of high-quality devices, and the longitudinal resistances $R_x(i)$ of this standard become very small over ranges of magnetic field in which quantized Hall resistance plateaus are observed. The dc longitudinal resistance is defined to be $R_x(i) = V_x(i)/I_T$, where $V_x(i)$ is the measured longitudinal voltage drop between potential probes located on the same side of the device. The dc values of $R_x(i)$ can also be temperature [13] and current [14] dependent.

4. Equivalent Electrical Circuit of an AC QHE Standard

The quantized Hall resistance $R_H(i)$ of an ac QHE resistance standard (ac QHRS) can be experimentally compared with the impedances of ac reference stan-

dards using ac measurement systems. NIST initially plans to use ac resistors as reference standards, and an ac ratio bridge measurement system for the comparisons.

Figure 1 shows an equivalent electrical circuit representation of an ac QHRS in which the QHRS is being measured with an ac bridge using four-terminal-pair [15,16] techniques. (Neither the ac reference standard nor the ac ratio bridge are shown in the figure.) This circuit of an ac QHRS is rather detailed, so we explain it one step at a time, starting with the periphery of the standard, then proceeding to the QHE device within the central region of the figure, and finally discussing properties of the sample probe leads within the standard.

The ac QHRS of Fig. 1 is bounded by an electrical shield indicated schematically by thick lines. Actual shields have complicated surface geometries. They consist of: (a) conductive surfaces surrounding the QHE device and its sample holder at liquid helium temperatures; (b) the outer conductors of eight coaxial leads within the sample probe; and (c) the outer conductors of eight coaxial leads extending from the top of the sample probe to room temperature access points S, 1 through 6, and D. The electrical shields will also be referred to in the text as “outer conductors”. To simplify the figure, we label only currents in the inner conductors.

The ac QHRS has electrical access at room temperature via four coaxial measurement ports labeled Inner/Outer, Detector, Potential, and Drive. These four ports are used in the four-terminal-pair measurements, where each coaxial port is referred to as a “terminal-pair”. The four coaxial ports are connected to room temperature access points S, 4, 3, and D in the figure.

The ideal four-terminal-pair measurement definition [15,16] of $R_H(i)$ is satisfied by the following three simultaneous conditions: (1) the current I_{Dr} at the Drive coaxial port is adjusted so that there are no currents in the inner or outer conductors of the Potential coaxial port, i.e., $I_{Pt} = 0$; (2) the potential difference is zero across the inner and outer conductors of the Detector coaxial port; and (3) there are no currents in the inner or outer conductors of the Detector coaxial port, i.e., $I_{Dt} = 0$.

It is implicit in the four-terminal-pair definition that each coaxial port is treated as a terminal-pair, and that the current in the inner conductor of every port is equal and opposite to the current in the outer conductor (the shield). Coaxial chokes [17] (located outside the ac quantized Hall resistance standard and therefore not shown in the figure) assure that this equal and opposite current condition is satisfied for each of the four terminal-pairs in the circuit. The current I_{Ot} exits the ac QHRS at the Inner/Outer port and enters the ac reference standard (not shown).

A “virtual” short has been drawn in Fig. 1 as a line between the shield and inner conductor at the Detector coaxial port to indicate four-terminal-pair condition number (2). We let the Detector potential be zero, i.e., $V_{Dt} = 0$. At bridge balance the ac quantized Hall voltage $V_H(i) = V_H(3,4) = V_{Pt}$ is defined as

$$V_{Pt} = [1 + \Delta_H] R_H(i) I_{Ot}, \quad (1)$$


where Δ_H is the correction factor to $R_H(i)$ to be determined in this analysis.

Next we describe the equivalent circuit model of the QHE device located in the central dashed-line region of Fig. 1. This model is based on that of Ricketts and Kemeny [18]. The device has contact pads that provide electrical access to the 2DEG at the source S' , the drain D' , and the potential pads 1' through 6'. Each contact pad is located at the end of an arm of the QHE device. Every arm in the equivalent circuit has an intrinsic resistor whose value is $R_H(i)/2$. We assume that the device is homogeneous, i.e., that the quantized Hall resistances $R_H(i)$ are all measured on plateau regions, that their values are the same on all the Hall potential probe sets, and that they are all measured at the same magnetic flux density value. $R_H(i)$ can, however, vary with temperature [13] and current [14].

While V_{Pt} has been observed to vary with frequency [4–9], it is not clear whether this is due to a frequency dependence of $R_H(i)$, of Δ_H , or of both $R_H(i)$ and Δ_H . Calculations of the intrinsic impedance of the 2DEG due to the internal Hall capacitance across the QHE device [19], however, predict a negligible frequency dependence of $R_H(i)$ itself, implying a frequency dependence of Δ_H arising from parasitic impedances in the ac QHRS. We therefore simplify the model, and assume that the dc values are appropriate for the $R_H(i)/2$ resistances in the figure.

The symbols r_a , r_b , r_c , and r_d in Fig. 1 represent real (in-phase) longitudinal resistances within the QHE device. Their measured dc values can vary with temperature [13] and current [14]. Sample probes normally used in dc QHE measurements have ten leads, with a pair of leads to the source contact pad S' and another pair to the drain contact pad D' . Only one lead of each pair carries the current, so the dc values of all four longitudinal resistances r_a , r_b , r_c , and r_d can be obtained using four-terminal measurements.

In order to reduce the heat load on the liquid helium, sample probes for the ac QHE usually have a single coaxial lead to each of the eight contact pads. Therefore only r_b and r_c can be determined directly via four-terminal-pair ac measurements. For example, a four-terminal-pair ac longitudinal resistance measurement of r_b could



be made by moving the Potential coaxial port from access position 3 to position 2 in Fig. 1, and measuring the ac longitudinal voltage $V_x(2,4)$.

$$V_x(2,4) = [1 + \Delta_{24}] r_b I_{0t}, \quad (2)$$

where Δ_{24} is the correction factor to r_b to be determined in this analysis. Values for r_a and r_d could be estimated from their dc r_a/r_b and r_d/r_c ratios if the measured r_b/r_c ratio happens to be the same for both ac and dc measurements using the same sample probe during the same cool-down.

With one exception [20], the reported ac longitudinal resistances obtained from the real, in-phase components of the ac longitudinal voltage measurements are significantly larger than the dc longitudinal resistances in the same device under the same temperature and magnetic field conditions. The ac longitudinal resistances increase with increasing frequency of the applied current, and are of order 1 m Ω at 1592 Hz [4,5,21]. The large ac longitudinal voltages might be due to intrinsic frequency dependences of r_a , r_b , r_c , and r_d within the device, to Δ_{24} , Δ_{46} , etc. corrections caused by parasitic impedances of the QHRS, or to both of them. Calculations of the kinetic inductance of the 2DEG and the magnetic inductance of the device [20] provide no plausible explanations via intrinsic impedance for significant frequency dependences of r_a , r_b , r_c , and r_d , suggesting that the frequency dependence of the ac longitudinal resistance is due to parasitic impedances of the QHRS, and therefore to the correction factors Δ_{24} , Δ_{46} , etc. However, we will assume the worst-case scenario in our numerical calculations, that is r_a , r_b , r_c , and r_d are themselves frequency dependent and have 1 m Ω values at 1592 Hz.

At some moment in time, a positive current I_a enters the 2DEG via device drain contact pad D' in Fig. 1, and current I_d exits the 2DEG via source contact pad S'. The magnetic flux density B is directed into the figure from above. Under these current and magnetic field conditions, the drain contact pad D' and the potential probe contact pads 1', 3', and 5' at the device periphery are at higher potentials than contact pads S', 2', 4', and 6'. These current and flux density directions are chosen to be consistent with those we have used in earlier calculations [19,22–23].

Potentials at the contact pads S', 1' through 6', and D' are produced by arrays of voltage generators, where each voltage generator V_{AB} is located between a pair of arms A and B of the equivalent circuit. The voltages are defined as

$$V_{AB} \equiv \frac{R_H(i)}{2} |I_A \pm I_B|, \quad (3)$$

where I_A and I_B are the magnitudes of the current flowing in arms A and B. The currents I_A and I_B within the absolute quantity sign of Eq. (3) are added if they both enter or both leave the voltage generator, and are subtracted if one current enters and the other current leaves the generator. For example $V_{1D} = [R_H(i)/2] |I_a - I_c|$. The voltages generated are functions of $R_H(i)$; therefore their values can vary with temperature [13] and current [14] (and also possibly with frequency).

Diamond-shaped voltage generator arrays of Ricketts and Kemeny [18] are employed in the equivalent circuit of the QHE device, rather than the ring-shaped voltage generator arrays introduced later by Delahaye [24] and then subsequently used by Jeffery, Elmquist, and Cage [25]. Although both arrays give essentially identical results [22], the calculations are much simpler with the diamond arrays when longitudinal resistances are included in the circuits [22]. We therefore use diamond arrays.

For clarity, the voltage generators are indicated in the figure as batteries, with positive terminals oriented to give the correct potentials along each arm at the instant considered. The ac currents alternate direction, so the voltage generators reverse sign each half cycle. Thus, for the part of the period in which the currents flow in the directions indicated in Fig. 1, the voltage generators have the polarities shown. Half a period later the currents change direction, and all the voltage generators reverse polarities.

The QHE device is mounted on a sample holder at the bottom of the sample probe. The QHE device and the sample holder are located within the shaded region of Fig. 1. Thin wires connect the device contact pads S', 1' through 6', and D' to coaxial leads which extend to room temperature access points S, 1 through 6, and D located outside the sample probe (but still within the ac QHRS). Each arm of the equivalent circuit has a resistance r_s , r_1 through r_6 , or r_D . This resistance includes the contact resistance to the 2DEG, the wire resistance connecting a contact pad on the device to a coaxial lead, and the inner conductor resistance of that coaxial lead. The inner conductor lead resistances vary with the liquid helium level in the sample probe. They can be measured pair-wise (using access points S, 1 through 6, and D) as a function of liquid helium level via two-terminal dc resistance measurements by temporarily replacing the QHE device with electrical shorts at positions S', 1' through 6', and D'. The cooled inner conductor coaxial lead resistances are typically each about 1 Ω in ac quantized Hall resistance standards. The outer conductor coaxial lead resistances depend on the type of coaxial cable, and their values also vary with liquid helium level. Typical values range between about 0.1 Ω and 1 Ω in ac quantized Hall resistance experiments.

Each sample probe lead has an inductance L_S , L_1 through L_6 , or L_D , that is electrically connected in series with the lead resistances r_S , r_1 through r_6 , or r_D , producing lead impedances z_S , z_1 through z_6 , or z_D , where $z_S = r_S + j\omega L_S$. Due to severe space limitations in the figure, these impedances are unconventionally drawn as resistors within rectangles. The inductance of each coaxial lead of a typical ac QHE sample probe is about 1×10^{-6} H. We assume that the bonding pad wires are thick enough to not vibrate in the magnetic field when applied ac currents flow through them [4], but the out-of-phase “inductance” generated by this vibration [4] could be included in the lead inductances if necessary.

The eight coaxial leads, labeled S, 1 through 6, and D, each have an inner and an outer conductor. The outer conductors of the coaxial leads are connected together outside the sample probe to help satisfy the four-terminal-pair measurement conditions. As mentioned earlier, the outer conductors of these leads act as electrical shields, and are represented schematically as thick lines in Fig. 1. (Other outer conductors of the ac QHRS also contribute to the thick lines.)

Large capacitances-to-shield, labeled as C_S , C_1 through C_6 , and C_D , exist between the inner and outer conductors of these coaxial leads. The open-circuit capacitances can be individually measured at access points S, 1 through 6, and at D as a function of liquid helium level by temporarily removing the QHE device from the sample probe at the points S', 1' through 6', and D'. The capacitance-to-shield of each coaxial lead in typical ac QHE sample probes is about 250 pF, but it should be reduced to about 100 pF (1×10^{-10} F) in a short sample probe being designed at NIST.

A predominately 90° out-of-phase current I_{C_S} , I_{C_1} through I_{C_6} , or I_{C_D} flows through each coaxial lead. These currents, and all the other currents in Fig. 1, have the correct signs for their dominant phase components in the half-cycle under consideration. This is verified in Sec. 5, where it is found that all currents shown in the figure have positive signs for their major components.

The coaxial leads are not the only sources of capacitances-to-shield. There are also additional contributions from the QHE device—sample holder combination and the electrical shielding surrounding them. These additional capacitances-to-shield are labeled C_A and C_B in Fig. 1, where they are placed at either end of the QHE device. (Note that rather than explicitly using C_A and C_B , one-eighth of the additional capacitances $C_A + C_B$ could instead be added to each of the eight coaxial lead capacitances C_S , C_1 through C_6 , and C_D , but that would make the coaxial lead capacitance notation very confusing.)

The additional capacitances C_A and C_B can be determined by two methods. In the first method the magnetic field is adjusted so the QHE device is on a QHE plateau.

The external coaxial leads from the bridge are removed from the Drive and Inner/Outer ports of the ac QHRS, and an applied voltage signal is placed across the inner and outer conductors of the Drive port. A measured voltage signal appears across the inner and outer conductors of coaxial leads S, D, 1, 3, and 5 for the magnetic field direction assumed in Fig. 1, so these particular coaxial leads draw most of the 90° out-of-phase current. Therefore the measured total capacitance-to-shield C_T is approximately $C_T(B) \approx C_1 + C_3 + C_5 + C_D + C_A$, and the value of C_A can be obtained by subtracting the value of $C_1 + C_3 + C_5 + C_D$ from $C_T(B)$. The magnetic field is reversed. Then $C_T(-B) \approx C_2 + C_4 + C_6 + C_S + C_B$ when the voltage signal is placed across the inner and outer conductors at the Inner/Outer port, thus yielding the value of C_B . In the second method the magnetic flux density B is reduced to zero. The quantum Hall voltages disappear, so the voltage generators can be replaced in the circuit by electrical shorts. The QHE device now behaves like a two-dimensional sheet resistance, and the $R_H(i)/2$ resistances located at the source and drain ends of the QHE device in Fig 1 are zero. Longitudinal resistances r_a , r_b , r_c , and r_d become much larger than they were when on a QHE plateau. Their values can be obtained by four-terminal resistance measurements in a dc sample probe. The $R_H(i)/2$ resistances of the six QHE side arms are replaced by much smaller resistances whose values can be obtained from two-terminal measurements via room temperature access points S, 1 through 6, and D once the appropriate lead and longitudinal resistances are subtracted. An applied voltage signal placed across the inner and outer conductors of the Drive port would cause a voltage signal to appear across the inner and outer conductors of all capacitances-to-shield. Thus the total capacitance-to-shield is given by the expression $C_T = C_S + C_1 + C_2 + C_3 + C_4 + C_5 + C_6 + C_D + C_A + C_B$, where $C_A \approx C_B$ if the QHE device, the sample holder, and the bonding wires between them are all symmetrically arranged. We expect both C_A and C_B will be about 1 pF or smaller in the NIST sample probes.

The equivalent circuit accounts for leakage currents between the ac QHRS's inner conductors and the shields via resistances r_{K_A} and r_{K_B} located on either side of the QHE device. Rather large voltages are used when measuring leakage resistances, so it would be safest to temporarily replace the device with shorts when measuring the total open-circuit leakage resistance r_{LK} at access point S, 1 through 6, or D. If the leakage resistances are symmetrically distributed, then $r_{K_A} \approx r_{K_B} \approx 2r_{LK}$. (Their values are large compared with the lead resistances, so they are essentially connected in parallel within the circuit.) The NIST sample probes will be constructed so these leakage resistances are very large; r_{K_A} and r_{K_B}

should be at least $10^{14} \Omega$, but in the numerical examples of this paper we will assume $10^{12} \Omega$ arising from dirty coaxial connectors.

The capacitances, inductances, and leakage resistances of Fig. 1 contribute parasitic impedances to measurements of the ac QHRS. These capacitances, inductances, and leakage resistances are drawn as discrete circuit elements. In reality they are distributed within the standard. They could, in principle, be better represented. For example, we could replace capacitance-to-shield C_1 with a capacitor of value $C_1/2$, and place a second capacitor of value $C_1/2$ and a series-connected outer shield impedance z'_1 between the other side of circuit element z_1 at point 1' and the first $C_1/2$ capacitor. This distributed impedance would, however, greatly complicate the circuit analyses, with little gain in accuracy. (Our discrete elements circuit over-emphasizes the capacitance-to-shield currents if $z_1 \geq z'_1$ and gives the same capacitance-to-shield currents if $z_1 = z'_1$.)

This completes the description of the equivalent circuit. The next section analyzes the circuit.

5. Analysis of the Single-Series “Normal” Circuit

Kirchoff's rules are used to sum the currents at branch points and the voltages around loops to obtain exact algebraic equations for the equivalent electrical circuit shown in Fig. 1. We refer to this circuit as single-series “normal”: single-series because there is just one current lead connected to the source contact pad S', and another current lead connected to the drain pad D' of the QHE device; and “normal” because the Hall voltage leads are connected to the central arms 3 and 4 of the device.

5.1 Exact Single-Series “Normal” Equations

Finding the exact algebraic equations for all the currents, and for the correction factor Δ_H to the quantum Hall voltage, as defined by Eq. (1), is rather difficult because there are many coupled equations, especially for the multi-series circuits [24] examined later in this paper. All the solutions of this paper were independently derived by each author, and shown to be identical. Then each author independently used computer software to obtain identical numerical results for several test cases. It is important to obtain the exact solutions, rather than initially guess approximate solutions, because the frequency dependent effects we are trying to minimize or eliminate are small, but significant. The results are presented here in order to spare others the task of deriving them.

To simplify the final algebraic expressions, we define some substitutions of variables, and substitutions of substitutions. The particular substitutions depend on the choice of loops. For example, the variables A and B listed below result from a voltage loop around the path C_S , S', C_B , and back through the shield to C_S . This gives the equation $-\frac{1}{j\omega C_S} I_{C_S} + z_S I_S + \frac{1}{j\omega C_B} I_{C_B} = 0$, so $I_{C_B} = \frac{C_B}{C_S} I_{C_S} - j\omega C_B z_S I_S$, or $I_{C_B} = A I_{C_S} - B I_S$. Let

$$A = \frac{C_B}{C_S} \quad (4a)$$

$$B = j\omega C_B z_S \quad (4b)$$

$$C = \frac{1}{j\omega C_B r_{KB}} \quad (4c)$$

$$D = \frac{1}{[1 + j\omega C_6(R_H + z_6)]} \frac{C_6}{C_B} \quad (4d)$$

$$E = \frac{j\omega C_6 r_d}{[1 + j\omega C_6(R_H + z_6)]} \quad (4e)$$

$$F = \frac{1}{[1 + j\omega C_5 z_5]} \frac{C_5}{C_B} \quad (4f)$$

$$G = \frac{j\omega C_5(R_H + r_d)}{[1 + j\omega C_5 z_5]} \quad (4g)$$

$$H = \frac{[1 + j\omega C_5(R_H + z_5)]}{[j\omega C_5(R_H - r_c)]} \quad (4h)$$

$$I = 1 + C + D + F(1 - H) \quad (4i)$$

$$J = E + G(1 - H) \quad (4j)$$

$$K = I + J(1 + C) \quad (4k)$$

$$L = \frac{AK}{[1 + J + K(A + B)]} \quad (4l)$$

$$M = \frac{j\omega C_3 R_H}{[1 + j\omega C_3 z_3]} \quad (4m)$$

$$N = \frac{j\omega C_2 r_b}{[1 + j\omega C_2(R_H + z_2)]} \quad (4n)$$

$$O = \frac{j\omega C_1(R_H + r_b)}{[1 + j\omega C_1 z_1]} \quad (4o)$$

$$P = \frac{(R_H + r_a)}{r_{KA}} \quad (4p)$$

$$Q = \frac{[1 + j\omega C_2 z_2]}{[j\omega C_2 r_{K_A}]} \quad (4q)$$

$$R = j\omega C_A r_{K_A} \quad (4r)$$

$$S = \frac{C_D}{C_A} \quad (4s)$$

$$T = j\omega C_D z_D. \quad (4t)$$

We express all currents, and the quantum Hall voltage, in terms of I_{Ot} because that is the current that enters the ac reference standard (not shown in Fig. 1). Three of the current solutions are trivial because of the four-terminal-pair definition [15,16] listed in Sec. 4

$$I_{Dr} = I_{Pt} = I_{C_4} = 0. \quad (5a)$$

The remaining exact equations for the single-series “normal” circuit currents are

$$I_S = LI_{Ot} \quad (5b)$$

$$I_{C_S} = I_{Ot} - I_S \quad (5c)$$

$$I_{C_B} = AI_{C_S} - BI_S \quad (5d)$$

$$I_{K_B} = CI_{C_B} \quad (5e)$$

$$I_d = I_S - I_{C_B} - I_{K_B} \quad (5f)$$

$$I_{C_6} = DI_{C_B} - EI_d \quad (5g)$$

$$I_{C_5} = -FI_{C_B} + GI_d \quad (5h)$$

$$I_c = HI_{C_5} \quad (5i)$$

$$I_{C_3} = MI_c \quad (5j)$$

$$I_b = I_c + I_{C_3} \quad (5k)$$

$$I_{C_2} = NI_b \quad (5l)$$

$$I_{C_1} = OI_b \quad (5m)$$

$$I_a = I_b + I_{C_1} + I_{C_2} \quad (5n)$$

$$I_{K_A} = PI_a + QI_{C_2} \quad (5o)$$

$$I_{C_A} = RI_{K_A} \quad (5p)$$

$$I_D = I_a + I_{K_A} + I_{C_A} \quad (5q)$$

$$I_{C_D} = SI_{C_A} + TI_D \quad (5r)$$

$$I_{Dr} = I_D + I_{C_D}. \quad (5s)$$

The exact equation for the quantum Hall voltage is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port. Taking the path through arm 4, voltage generators V_{c4} and V_{c3} , and arm 3 we find that

$$V_H(3,4) = R_H I_c - z_3 I_{C_3}, \quad (6a)$$

which can also be expressed in the form

$$V_H(3,4) = [1 + \Delta_{34}] R_H I_{Ot} \quad (6b)$$

by using Eqs. (4) and (5). An approximate solution in this form will be given in Eq. (11).

5.2 A Numerical Example

Contributions of the parasitic impedance within the ac QHRS to the measured value of $V_H(3,4)$ can be investigated by using numerical examples in Eqs. (5) and (6). We assign cardinal numbers to circuit element values to emphasize that the results are not intended to provide corrections to existing experimental data because the effects of wire-to-wire capacitances are not included at this intermediate stage of the analysis.

Both the $i = 2$ (12 906.4 Ω) and $i = 4$ (6 453.2 Ω) plateaus have been measured in ac experiments, so let $R_H = 10\,000\, \Omega$. The cardinal values we use are

$$R_H = 10^4\, \Omega \quad (7a)$$

$$r_S = r_1 = r_2 = r_3 = r_4 = r_5 = r_6 = r_D = 1\, \Omega \quad (7b)$$

$$r_a = r_b = r_c = r_d = 10^{-3}\, \Omega \quad (7c)$$

$$r_{K_A} = r_{K_B} = 10^{12}\, \Omega \quad (7d)$$

$$C_S = C_1 = C_2 = C_3 = C_4 = C_5 = C_6 = C_D = 10^{-10}\, \text{F} \quad (7e)$$

$$C_A = C_B = 10^{-12}\, \text{F} \quad (7f)$$

$$L_S = L_1 = L_2 = L_3 = L_4 = L_5 = L_6 = L_D = 10^{-6}\, \text{H} \quad (7g)$$

$$\omega = 10^4\, \text{rad/s}. \quad (7h)$$

Note that the 100 pF capacitances-to-shield values of Eq. (7e) may be close to those that will be obtained in the short NIST sample probe, but typical ac probes have values around 250 pF.

The numerical results for the currents of Eqs. (5) are

$$I_{C_S} = \{-[1.0 \times 10^{-8}] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (8a)$$

$$I_S = \{[1.00000] - j[1.0 \times 10^{-6}]\}I_{Ot} \quad (8b)$$

$$I_{C_B} = \{-[2.0 \times 10^{-13}] + j[2.0 \times 10^{-11}]\}I_{Ot} \quad (8c)$$

$$I_{K_B} = \{[2.0 \times 10^{-15}] + j[2.0 \times 10^{-17}]\}I_{Ot} \quad (8d)$$

$$I_d = \{[1.00000] - j[1.0 \times 10^{-6}]\}I_{Ot} \quad (8e)$$

$$I_{C_6} = \{-[1.0 \times 10^{-11}] + j[1.0 \times 10^{-9}]\}I_{Ot} \quad (8f)$$

$$I_{C_5} = \{[2.0 \times 10^{-8}] + j[0.01000]\}I_{Ot} \quad (8g)$$

$$I_c = \{[1.00000] + j[0.01000]\}I_{Ot} \quad (8h)$$

$$I_{C_3} = \{-[1.0 \times 10^{-4}] + j[0.01000]\}I_{Ot} \quad (8i)$$

$$I_b = \{[0.99990] + j[0.02000]\}I_{Ot} \quad (8j)$$

$$I_{C_2} = \{-[1.0 \times 10^{-11}] + j[1.0 \times 10^{-9}]\}I_{Ot} \quad (8k)$$

$$I_{C_1} = \{-[2.0 \times 10^{-4}] + j[0.01000]\}I_{Ot} \quad (8l)$$

$$I_a = \{[0.99970] + j[0.03000]\}I_{Ot} \quad (8m)$$

$$I_{K_A} = \{[1.0 \times 10^{-8}] + j[3.0 \times 10^{-10}]\}I_{Ot} \quad (8n)$$

$$I_{C_A} = \{-[3.0 \times 10^{-6}] + j[1.0 \times 10^{-4}]\}I_{Ot} \quad (8o)$$

$$I_D = \{[0.99970] + j[0.03010]\}I_{Ot} \quad (8p)$$

$$I_{C_D} = \{-[3.0 \times 10^{-4}] + j[0.01000]\}I_{Ot} \quad (8q)$$

$$I_{D_r} = \{[0.99940] + j[0.04010]\}I_{Ot} \quad (8r)$$

The 90° out-of-phase (j) parts of shunt currents I_{C_5} , I_{C_3} , I_{C_1} , I_{C_D} , and I_{C_A} , are much larger than for shunt currents I_{C_2} , I_{C_6} , I_{C_S} , and I_{C_B} because contact pads 5', 3', 1', and D' are all near the quantum Hall potential, rather than near the shield potential. A 1 % out-of-phase current passes through each of the coaxial cable capacitances C_5 , C_3 , C_1 , and C_D in this example. That is not necessarily a problem if the bridge Drive can provide this extra 4 % of out-of-phase current to I_{D_r} .

Expressing Eq. (6a) in the form of Eq. (6b), we find that

$$V_H(3,4) = \{1 + [5.0 \times 10^{-8}] + j[0.01000]\}R_H I_{Ot}, \quad (9a)$$

$$\Delta_{34} = \{[5.0 \times 10^{-8}] + j[0.01000]\}. \quad (9b)$$

The 5×10^{-8} in-phase correction to R_H is too large compared with the desired 1×10^{-8} R_H absolute accuracy, but even worse, there is a 1 % contribution to $V_H(3,4)$ in the 90° out-of-phase j term. Auxiliary balances in the

NIST high precision ac bridges are not capable of providing out-of-phase adjustment signals larger than $5 \times 10^{-4} R_H$, so the 1 % out-of-phase signal is unacceptable. We next list the approximate solutions to show the source of this problem.

5.3 Approximate Single-Series “Normal” Solutions

Many of the terms in the following approximate solutions were obtained by algebraically finding the dominant contributions to the exact equations. Other terms were found by “educated guesses” and “trial-and-error”. We verified all the terms by changing the values of relevant circuit element components in the computer programs. The following approximate solutions give numerical results that agree with the results from the exact solutions to within at least two significant figures for both the real and imaginary parts of the numerical results. Other terms may need to be added to these approximate equations if the circuit components have values significantly different from those listed in Eqs. (7).

$$I_{C_S} \approx \{-[\omega^2 C_S(C_5 + C_6)R_H r_c - \omega^2 C_S C_S r_S r_S + \omega^2 C_S L_S] + j[\omega C_S r_S]\}I_{Ot} \quad (10a)$$

$$I_S \approx \{1 - j[\omega C_S r_S]\}I_{Ot} \quad (10b)$$

$$I_{C_B} \approx \{-[\omega^2 C_B(C_5 + C_6)R_H r_c] + j[\omega C_B(r_c + r_d)]\}I_{Ot} \quad (10c)$$

$$I_{K_B} \approx \left\{ \left[\frac{(r_c + r_d)}{r_{K_B}} \right] + j \left[\omega(C_5 + C_6) \frac{R_H r_c}{r_{K_B}} \right] \right\} I_{Ot} \quad (10d)$$

$$I_d \approx \{1 - j[\omega C_S r_S]\}I_{Ot} \quad (10e)$$

$$I_{C_6} \approx \{-[\omega^2 C_5 C_6 R_H r_c] + j[\omega C_6 r_c]\}I_{Ot} \quad (10f)$$

$$I_{C_5} \approx \{[\omega^2 C_S C_5 R_H r_S + \omega^2 C_5 C_5 R_H r_S] + j[\omega C_5 R_H]\}I_{Ot} \quad (10g)$$

$$I_c \approx I_{Ot} + I_{C_5} - I_{C_S} \quad (10h)$$

$$I_{C_3} \approx \{-[\omega^2 C_3 C_5 R_H R_H] + j[\omega C_3 R_H]\}I_{Ot} \quad (10i)$$

$$I_b \approx I_{Ot} + I_{C_5} + I_{C_3} \quad (10j)$$

$$I_{C_2} \approx \{-[\omega^2 C_2(C_3 + C_5 - C_2)R_H r_b] + j[\omega C_2 r_b]\}I_{Ot} \quad (10k)$$

$$I_{C_1} \approx \{-[\omega^2 C_1(C_3 + C_5)R_H R_H] + j[\omega C_1 R_H]\}I_{Ot} \quad (10l)$$

$$I_a \approx I_{Ot} + I_{C_5} + I_{C_3} + I_{C_1} \quad (10m)$$

$$I_{K_A} \approx \left\{ \left[\frac{R_H}{r_{K_A}} \right] + j \left[\omega(C_1 + C_3 + C_5) \frac{R_H R_H}{r_{K_A}} \right] \right\} I_{Ot} \quad (10n)$$

$$I_{C_A} \approx \{ -[\omega^2 C_A (C_1 + C_3 + C_5) R_H R_H] + j[\omega C_A R_H] \} I_{O_t} \quad (10o)$$

$$I_D \approx I_{O_t} + I_{C_5} + I_{C_3} + I_{C_1} + I_{C_A} \quad (10p)$$

$$I_{C_D} \approx \{ -[\omega^2 C_D (C_1 + C_3 + C_5) R_H R_H] + j[\omega C_D R_H] \} I_{O_t} \quad (10q)$$

$$I_{D_r} \approx I_{O_t} + I_{C_5} + I_{C_3} + I_{C_1} + I_{C_A} + I_{C_D}. \quad (10r)$$

Expressing Eq. (6a) in the form (6b),

$$\begin{aligned} \Delta_H \approx & \{ [\omega^2 C_S C_5 R_H r_S + \omega^2 C_3 C_5 R_H r_3 + \omega^2 C_5 C_5 R_H r_5 \\ & - \omega^2 C_S C_5 r_S r_S] + [\omega^2 C_S (C_5 + C_6) R_H r_c + \omega^2 C_S L_S \\ & + \omega^2 C_3 L_3] + j[\omega C_5 R_H - \omega C_S r_S - \omega C_3 r_3] \}. \end{aligned} \quad (11)$$

We see from Eq. (11) that sample probe lead 5 is the dominant source of the 1 % out-of-phase component of the quantum Hall voltage signal. The next subsection investigates the effect of removing this lead.

5.4 Disconnecting Sample Probe Lead 5

Equation (11) predicts that the out-of-phase term $j[\omega C_5 R_H]$ in the expression for Δ_{34} can be reduced by disconnecting coaxial cable 5 at position 5', where 5' is either located at the potential contact pad on the QHE device, or at an intermediate contact point in the sample holder. There is a capacitance $C_{5'}$ between the QHE device and the shield that replaces capacitance C_5 in Fig. 1. Also, a shield impedance $z_{5'}$ replaces the lead impedance z_5 .

The most significant terms of Eq. (11) are now

$$\begin{aligned} \Delta_{34} \approx & \{ [\omega^2 C_S L_S + \omega^2 C_3 L_3] + \\ & j[\omega C_{5'} R_H - \omega C_S r_S - \omega C_3 r_3] \}. \end{aligned} \quad (12)$$

If we assume in the numerical examples that

$$C_{5'} = C_A = C_B = 1 \text{ pF} \quad (13)$$

$$r_{5'} = r_5 = 1 \text{ } \Omega. \quad (14)$$

Then

$$\Delta_{34} = \{ [2.0 \times 10^{-8}] + j[9.8 \times 10^{-5}] \} \quad (15)$$

when the coaxial lead capacitances are all 100 pF, and

$$\Delta_{34} = \{ [5.0 \times 10^{-8}] + j[9.5 \times 10^{-5}] \} \quad (16)$$

when they are 250 pF. All experiments which have measured ac values of $V_H(3,4)$ have had to remove coaxial

lead 5 because of the effects due to the large capacitance-to-shield C_5 presented above.

Equations (11) or (12) might be used to apply corrections to the experimental data in order to reduce the 5×10^{-8} in-phase error in $R_H I_{O_t}$. However, there are several points of concern: (a) our approximate and exact equations do not include the effects of wire-to-wire capacitances, and these may be significant; (b) the out-of-phase component of $V_H(3,4)$ has been reduced to about $1 \times 10^{-4} R_H I_{O_t}$ by removing lead 5, but great care must be taken to correct for the in-phase (phase defect) contributions of the bridge components used to null the out-of-phase signal because these in-phase (phase defect) signals can be unintentionally added to the real, second-order terms of the in-phase component of $V_H(3,4)$ in Eqs. (11) or (12) that vary with ω^2 ; and (c) it is not trivial to measure the value of C_5 in order to apply the correction with lead 5 disconnected.

We will not consider further the single-series “normal” circuit as a viable ac QHRS candidate because lead 5 must be disconnected, and that violates one of our desired goals.

6. Analysis of the Single-Series “Offset” Circuit

Figure 2 shows an equivalent electrical circuit representation of an ac QHRS using single-series “offset” connections to the QHE device. It is single-series because there is just one current lead connected to the source contact pad S', and another current lead connected to the drain pad D' of the QHE device, and “offset” because the Hall voltage leads are connected to the off-center arms 5 and 6 of the device. Arms 5 and 6 are closest to the low potential end of the device at S', and nearest to the ac reference resistor (not shown in the figure). Those arms were chosen in an attempt to reduce the effects of shunt currents through I_{C_5} that we found in Sec. 5.

6.1 Exact Single-Series “Offset” Equations

To simplify the final algebraic expressions, we again define some intermediate substitutions of variables, and substitutions of substitutions. Let

$$A = \frac{C_B}{C_S} \quad (17a)$$

$$B = j\omega C_B z_S \quad (17b)$$

$$C = \frac{1}{j\omega C_B r_{K_B}} \quad (17c)$$




Fig. 2. An equivalent electrical circuit representation of an ac QHE resistance standard with single-series "offset" connections to the device. See Sec. 6 for the circuit analysis.

$$D = j\omega C_s z_s \quad (17d)$$

$$E = j\omega C_s r_d \quad (17e)$$

$$F = BE(1 + C) \quad (17f)$$

$$G = AE(1 + C) \quad (17g)$$

$$H = 1 + D + E + F + G \quad (17h)$$

$$I = \frac{j\omega C_5 R_H}{[1 + j\omega C_5 z_5]} \quad (17i)$$

$$J = \frac{j\omega C_4 r_c}{[1 + j\omega C_4 (R_H + z_4)]} \quad (17j)$$

$$K = \frac{j\omega C_3 (R_H + r_c)}{[1 + j\omega C_3 z_3]} \quad (17k)$$

$$L = \frac{j\omega C_2 r_b}{[1 + j\omega C_2 (R_H + z_2)]} \quad (17l)$$

$$M = \frac{\left[\frac{C_2}{C_4} + j\omega C_2 z_4 \right]}{[1 + j\omega C_2 (R_H + z_2)]} \quad (17m)$$

$$N = \frac{j\omega C_1 r_b}{[1 + j\omega C_1 z_1]} \quad (17n)$$

$$O = \frac{\left[\frac{C_1}{C_3} + j\omega C_1 (R_H + z_3) \right]}{[1 + j\omega C_1 z_1]} \quad (17o)$$

$$P = \frac{(R_H + r_a)}{r_{K_A}} \quad (17p)$$

$$Q = \frac{[1 + j\omega C_2 z_2]}{[j\omega C_2 r_{K_A}]} \quad (17q)$$

$$R = j\omega C_A r_{K_A} \quad (17r)$$

$$S = \frac{C_D}{C_A} \quad (17s)$$

$$T = j\omega C_D z_D. \quad (17t)$$

Three of the current solutions are trivial because of the four-terminal-pair definition [15,16]

$$I_{Dt} = I_{Pt} = I_{C_6} = 0. \quad (18a)$$

The remaining exact equations for the single-series “offset” circuit currents are

$$I_S = \frac{(1 + G)}{H} I_{Ot} \quad (18b)$$

$$I_{C_5} = I_{Ot} - I_S \quad (18c)$$

$$I_{C_B} = AI_{C_5} - BI_S \quad (18d)$$

$$I_{K_B} = CI_{C_B} \quad (18e)$$

$$I_d = I_S - I_{C_B} - I_{K_B} \quad (18f)$$

$$I_{C_5} = II_d \quad (18g)$$

$$I_c = I_d + I_{C_5} \quad (18h)$$

$$I_{C_4} = JI_c \quad (18i)$$

$$I_{C_3} = KI_c \quad (18j)$$

$$I_b = I_c + I_{C_3} + I_{C_4} \quad (18k)$$

$$I_{C_2} = LI_b + MI_{C_4} \quad (18l)$$

$$I_{C_1} = NI_b + OI_{C_3} \quad (18m)$$

$$I_a = I_b + I_{C_1} + I_{C_2} \quad (18n)$$

$$I_{K_A} = PI_a + QI_{C_2} \quad (18o)$$

$$I_{C_A} = RI_{K_A} \quad (18p)$$

$$I_D = I_a + I_{K_A} + I_{C_A} \quad (18q)$$

$$I_{C_D} = SI_{C_A} + TI_D \quad (18r)$$

$$I_{Dr} = I_D + I_{C_D}. \quad (18s)$$

The exact equation for the quantum Hall voltage is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port. Taking the path through arm 6, voltage generators V_{S6} and V_{S5} , and arm 5 we find that

$$V_H(5,6) = R_H I_d - z_5 I_{C_5}, \quad (19a)$$

which can also be expressed in the form

$$V_H(5,6) = [1 + \Delta_{S6}] R_H I_{Ot}. \quad (19b)$$

6.2 A Numerical Example

We investigate the parasitic impedance contributions of the ac QHRS on the measured value of $V_H(5,6)$ by using the cardinal numbers listed in Eqs. (7) in Eqs. (18) and (19). The numerical results for the currents are

$$I_{C_5} = \{-[1.0 \times 10^{-8}] + j[1.0 \times 10^{-6}]\}I_{Ot} \quad (20a)$$

$$I_S = \{[1.00000] - j[1.0 \times 10^{-6}]\}I_{Ot} \quad (20b)$$

$$I_{C_B} = \{[1.0 \times 10^{-17}] + j[1.0 \times 10^{-11}]\}I_{Ot} \quad (20c)$$

$$I_{K_B} = \{[1.0 \times 10^{-15}] - j[1.0 \times 10^{-21}]\}I_{Ot} \quad (20d)$$

$$I_d = \{[1.00000] - j[1.0 \times 10^{-6}]\}I_{Ot} \quad (20e)$$

$$I_{C_5} = \{[2.0 \times 10^{-8}] + j[0.01000]\}I_{Ot} \quad (20f)$$

$$I_c = \{[1.00000] + j[0.01000]\}I_{Ot} \quad (20g)$$

$$I_{C_4} = \{[2.0 \times 10^{-15}] + j[1.0 \times 10^{-9}]\}I_{Ot} \quad (20h)$$

$$I_{C_3} = \{-[1.0 \times 10^{-4}] + j[0.01000]\}I_{Ot} \quad (20i)$$

$$I_b = \{[0.99990] + j[0.02000]\}I_{Ot} \quad (20j)$$

$$I_{C_2} = \{[3.0 \times 10^{-15}] + j[2.0 \times 10^{-9}]\}I_{Ot} \quad (20k)$$

$$I_{C_1} = \{-[2.0 \times 10^{-4}] + j[0.01000]\}I_{Ot} \quad (20l)$$

$$I_a = \{[0.99970] + j[0.03000]\}I_{Ot} \quad (20m)$$

$$I_{K_A} = \{[1.0 \times 10^{-8}] + j[3.0 \times 10^{-10}]\}I_{Ot} \quad (20n)$$

$$I_{C_A} = \{-[3.0 \times 10^{-6}] + j[1.0 \times 10^{-4}]\}I_{Ot} \quad (20o)$$

$$I_D = \{[0.99970] + j[0.03010]\}I_{Ot} \quad (20p)$$

$$I_{C_D} = \{-[3.0 \times 10^{-4}] + j[0.01000]\}I_{Ot} \quad (20q)$$

$$I_{D_r} = \{[0.99940] + j[0.04010]\}I_{Ot} \quad (20r)$$

The 90° out-of-phase parts of shunt currents I_{C_5} , I_{C_3} , I_{C_1} , I_{C_D} , and I_{C_A} are again much larger than for shunt currents I_{C_2} , I_{C_4} , I_{C_S} , and I_{C_B} because contact pads 5', 3', 1', and D' are all near the quantum Hall potential, rather than near the shield potential. A 1 % out-of-phase current once again passes through each of the coaxial cable capacitances C_5 , C_3 , C_1 , and C_D in this example, which is not necessarily a problem if the bridge Drive can provide this extra 4 % of out-of-phase current to I_{D_r} .

Expressing Eq. (19a) in the form of Eq. (19b), we find that

$$V_H(5,6) = \{1 + [2.0 \times 10^{-8}] - j[2.0 \times 10^{-6}]\}R_H I_{Ot}, \quad (21a)$$

$$\Delta_{56} = \{[2.0 \times 10^{-8}] - j[2.0 \times 10^{-6}]\} \quad (21b)$$

for 100 pF lead capacitances and

$$\Delta_{56} = \{[5.0 \times 10^{-8}] - j[5.0 \times 10^{-6}]\} \quad (22)$$

for 250 pF coaxial leads.

The 2×10^{-8} in-phase correction to R_H for 100 pF leads is larger than our desired 1×10^{-8} R_H total uncertainty, but a correction could be made to the measurements via the approximate equation

$$\Delta_{56} \approx \{[\omega^2 C_S L_S + \omega^2 C_S L_S] - j[\omega C_S r_S + \omega C_S r_S]\} \quad (23)$$

that might provide sufficient accuracy. We will therefore consider the single-series “offset” circuit as a possible ac QHRS in a future paper which includes the effects of wire-to-wire capacitances. The approximate equations for the currents will be given in that paper.

7. Analysis of the Double-Series Circuit

Figure 3 shows an equivalent electrical circuit representation of an ac QHRS using two double-series connections to the QHE device. It is called double-series because there are two current paths to the device provided by a short coaxial lead outside the sample probe that connects room temperature access points 3 and D at point Y. Another short coaxial lead connects access points 4 and S at point Z. Short coaxial leads connect point Y with the Drive and Potential ports, and point Z with the Inner/Outer and Detector ports. For simplicity, we have placed all the parasitic impedances of the short coaxial cables in the cables and coaxial connectors labeled Ot, Dt, Pt, and Dr. These connections were first used by Delahaye [24] in ac quantized Hall resistance measurements (but points Y and Z were at the sample holder rather than outside the cryostat). Most subsequent ac experiments have used double-series or triple-series connections.

7.1 Exact Double-Series Equations

To simplify the final algebraic expressions, we again define substitutions of variables, and substitutions of substitutions. Let

$$A = j\omega C_{Ot} r_{Ot} \quad (24a)$$

$$B = j\omega C_B z_S \quad (24b)$$

$$C = \frac{1}{j\omega C_B r_{K_B}} \quad (24c)$$

$$D = \frac{1}{[1 + j\omega C_6(R_H + z_6)]} \frac{C_6}{C_B} \quad (24d)$$

$$E = \frac{j\omega C_6 r_d}{[1 + j\omega C_6(R_H + z_6)]} \quad (24e)$$



$$F = \frac{1}{[1 + j\omega C_5 z_5]} \frac{C_5}{C_B} \quad (24f)$$

$$G = \frac{j\omega C_5(R_H + r_d)}{[1 + j\omega C_5 z_5]} \quad (24g)$$

$$H_1 = \frac{z_5}{(R_H + z_4)} \quad (24h)$$

$$H_2 = \frac{r_d}{(R_H + z_4)} \quad (24i)$$

$$H_3 = \frac{r_c}{(R_H + z_4)} \quad (24j)$$

$$H_4 = \frac{R_H}{(R_H + z_4)} \quad (24k)$$

$$I = H_2 + H_3(1 + G) + E(H_3 - H_4) \quad (24l)$$

$$J = I(1 + C) + FH_3 + D(H_3 - H_4) \quad (24m)$$

$$K = \frac{1}{[1 + H_1 + I + BJ]} \quad (24n)$$

$$L = j\omega C_3 r_c \quad (24o)$$

$$M_1 = \frac{C_3}{C_5} + j\omega C_3(R_H + z_5) \quad (24p)$$

$$M_2 = j\omega C_3 z_3 \quad (24q)$$

$$N_1 = \frac{j\omega C_2 r_b}{[1 + j\omega C_2(R_H + z_2)]} \quad (24r)$$

$$N_2 = \frac{j\omega C_2 z_4}{[1 + j\omega C_2(R_H + z_2)]} \quad (24s)$$

$$O_1 = \frac{j\omega C_1 r_b}{[1 + j\omega C_1 z_1]} \quad (24t)$$

$$O_2 = \frac{j\omega C_1(R_H + z_3)}{[1 + j\omega C_1 z_1]} \quad (24u)$$

$$O_3 = \frac{C_1}{C_3[1 + j\omega C_1 z_1]} \quad (24v)$$

$$P_1 = \frac{(R_H + r_a)}{r_{K_A}} \quad (24w)$$

$$P_2 = \frac{z_2}{r_{K_A}} + \frac{1}{j\omega C_2 r_{K_A}} \quad (24x)$$

$$Q_1 = j\omega C_A r_{K_A} \quad (24y)$$

$$Q_2 = r_a + z_D[1 + P_1(1 + Q_1)] \quad (24z)$$

$$Q_3 = z_D P_2(1 + Q_1) \quad (25a)$$

$$Q_4 = r_b + Q_2 + O_1(R_H + Q) \quad (25b)$$

$$Q_5 = Q_4 + N_1(Q_2 + Q_3) \quad (25c)$$

$$Q_6 = R_H + z_3 + Q_5 + O_2(R_H + Q_2) - M_2 O_3(R_H + Q_2) \quad (25d)$$

$$Q_7 = \frac{[Q_5 + L O_3(R_H + Q_2)]}{Q_6} \quad (25e)$$

$$Q_8 = \frac{[Q_5 + N_2(Q_2 + Q_3)]}{Q_6} \quad (25f)$$

$$Q_9 = \frac{M_1 O_3(R_H + Q_2)}{Q_6} \quad (25g)$$

$$R = j\omega C_D z_D \quad (25h)$$

$$S = \frac{C_D}{C_A} \quad (25i)$$

$$T = \frac{C_{Pt}}{C_D} \frac{1}{[1 + j\omega C_{Pt} r_{Pt}]} \quad (25j)$$

$$U_1 = \frac{C_{Dr}}{C_D} \quad (25k)$$

$$U_2 = j\omega C_{Dr} r_{Dr} \quad (25l)$$

Six of the current solutions are trivial because of the four-terminal-pair definition [15,16]

$$I_{Dt} = I_{Pt} = I_{Cs} = I_{CDt} = I_{C4} = I_{rDt} = 0. \quad (26a)$$

The remaining exact equations for the double-series circuit currents are

$$I_{C_{Ot}} = \frac{A}{(1 + A)} I_{Ot} \quad (26b)$$

$$I_{r_{Ot}} = I_{Ot} - I_{C_{Ot}} \quad (26c)$$

$$I_S = K I_{r_{Ot}} \quad (26d)$$

$$I_4 = I_{r_{Ot}} - I_S \quad (26e)$$

$$I_{C_B} = B I_S \quad (26f)$$

$$I_{K_B} = C I_{C_B} \quad (26g)$$

$$I_d = I_S + I_{C_B} + I_{K_B} \quad (26h)$$

$$I_{C_6} = DI_{C_B} + EI_d \quad (26i)$$

$$I_{C_5} = FI_{C_B} + GI_d \quad (26j)$$

$$I_c = I_d + I_{C_5} + I_{C_6} \quad (26k)$$

$$I_{3'} = Q_7 I_c + Q_8 I_4 + Q_9 I_{C_5} \quad (26l)$$

$$I_{C_3} = LI_c + M_1 I_{C_5} + M_2 I_{3'} \quad (26m)$$

$$I_3 = I_{3'} + I_{C_3} \quad (26n)$$

$$I_b = I_c + I_4 - I_{3'} \quad (26o)$$

$$I_{C_2} = N_1 I_b + N_2 I_4 \quad (26p)$$

$$I_{C_1} = O_1 I_b - O_2 I_{3'} + O_3 I_{C_3} \quad (26q)$$

$$I_a = I_b + I_{C_1} + I_{C_2} \quad (26r)$$

$$I_{K_A} = P_1 I_a + P_2 I_{C_2} \quad (26s)$$

$$I_{C_A} = Q_1 I_{K_A} \quad (26t)$$

$$I_D = I_a + I_{K_A} + I_{C_A} \quad (26u)$$

$$I_{C_D} = SI_{C_A} + RI_D \quad (26v)$$

$$I_{C_{Pt}} = TI_{C_D} \quad (26w)$$

$$I_{D_r} = I_D + I_3 + I_{C_D} + I_{C_{Pt}} \quad (26x)$$

$$I_{C_{Dr}} = U_1 I_{C_D} + U_2 I_{D_r} \quad (26y)$$

$$I_{Dr} = I_{D_r} + I_{C_{Dr}} \quad (26z)$$

The exact equation for the quantum Hall voltage is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port. Taking the path through point Z, arm 4, voltage generators V_{c4} and V_{c3} , arm 3, and point Y we find that

$$V_H(Y, Z) = R_H I_c + (R_H + z_4) I_4 + z_3 I_{3'} - r_{Pt} I_{C_{Pt}} \quad (27a)$$

which can also be expressed as

$$V_H(Y, Z) = [1 + \Delta_{YZ}] R_H I_{Ot}. \quad (27b)$$

7.2 A Numerical Example

We investigate the parasitic impedance contributions of the ac QHRS on the measured value of $V_H(Y, Z)$ for a particular example of the double-series circuit by using

the cardinal numbers listed in Eqs. (7), plus the following cardinal numbers for the additional circuit elements

$$r_{Ot} = r_{Dt} = r_{Pt} = r_{Dr} = 10^{-3} \Omega \quad (28a)$$

$$C_{Ot} = C_{Dt} = C_{Pt} = C_{Dr} = 10^{-12} \text{ F.} \quad (28b)$$

The numerical results for the currents in Eqs. (26) are

$$I_{C_{Ot}} = \{[1.0 \times 10^{-22}] + j[1.0 \times 10^{-11}]\} I_{Ot} \quad (29a)$$

$$I_{r_{Ot}} = \{[1.00000] - j[1.0 \times 10^{-11}]\} I_{Ot} \quad (29b)$$

$$I_s = \{[0.99990] + j[9.0 \times 10^{-11}]\} I_{Ot} \quad (29c)$$

$$I_{C_B} = \{-[1.0 \times 10^{-10}] + j[1.0 \times 10^{-8}]\} I_{Ot} \quad (29d)$$

$$I_{K_B} = \{[1.0 \times 10^{-12}] + j[1.0 \times 10^{-14}]\} I_{Ot} \quad (29e)$$

$$I_d = \{[0.99990] + j[1.0 \times 10^{-8}]\} I_{Ot} \quad (29f)$$

$$I_{C_6} = \{[1.1 \times 10^{-11}] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (29g)$$

$$I_{C_5} = \{-[1.0 \times 10^{-10}] + j[0.01000]\} I_{Ot} \quad (29h)$$

$$I_c = \{[0.99990] + j[0.01000]\} I_{Ot} \quad (29i)$$

$$I_4 = \{[1.0 \times 10^{-4}] - j[1.0 \times 10^{-10}]\} I_{Ot} \quad (29j)$$

$$I_{3'} = \{[1.0 \times 10^{-4}] + j[0.01000]\} I_{Ot} \quad (29k)$$

$$I_{C_3} = \{-[1.0 \times 10^{-4}] + j[0.01000]\} I_{Ot} \quad (29l)$$

$$I_3 = \{[1.6 \times 10^{-7}] + j[0.02000]\} I_{Ot} \quad (29m)$$

$$I_b = \{[0.99990] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (29n)$$

$$I_{C_2} = \{[1.0 \times 10^{-11}] + j[1.1 \times 10^{-9}]\} I_{Ot} \quad (29o)$$

$$I_{C_1} = \{-[1.0 \times 10^{-12}] + j[0.01000]\} I_{Ot} \quad (29p)$$

$$I_a = \{[0.99990] + j[0.01000]\} I_{Ot} \quad (29q)$$

$$I_{K_A} = \{[1.0 \times 10^{-8}] + j[1.0 \times 10^{-10}]\} I_{Ot} \quad (29r)$$

$$I_{C_A} = \{-[1.0 \times 10^{-6}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (29s)$$

$$I_D = \{[0.99990] + j[0.01010]\} I_{Ot} \quad (29t)$$

$$I_{C_D} = \{-[1.0 \times 10^{-4}] + j[0.01000]\} I_{Ot} \quad (29u)$$

$$I_{C_{Pt}} = \{-[1.0 \times 10^{-6}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (29v)$$

$$I_{D_r} = \{[0.99980] + j[0.04020]\} I_{Ot} \quad (29w)$$

$$I_{C_{Dr}} = \{-[1.0 \times 10^{-6}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (29x)$$

$$I_{Dr} = \{[0.99980] + j[0.04030]\} I_{Ot} \quad (29y)$$

The 90° out-of-phase parts of shunt currents I_{C_5} , I_{C_3} , I_{C_1} , I_{C_D} , I_{C_A} , $I_{C_{P1}}$, and $I_{C_{Dr}}$ are again much larger than for shunt currents I_{C_2} , I_{C_6} , I_{C_B} , and $I_{C_{Ot}}$ because contact pads 5', 3', 1', and D' are all near the quantum Hall potential, rather than near the shield potential. A 1 % out-of-phase current passes through each of the coaxial cable capacitances C_5 , C_3 , C_1 , and C_D in this example, which once again is not necessarily a problem if the bridge Drive can provide this extra 4 % of out-of-phase current to I_{Dr} .

Expressing Eq. (27a) in the form of Eq. (27b), we find that

$$V_H(Y,Z) = \{1 + [9.8 \times 10^{-9}] + j[0.01000]\} R_H I_{Ot} \quad (30a)$$

$$\Delta_{YZ} = \{[9.8 \times 10^{-9}] + j[0.01000]\} \quad (30b)$$

for 100 pF lead capacitances and

$$\Delta_{YZ} = \{[3.2 \times 10^{-8}] + j[0.02501]\} \quad (31)$$

for 250 pF coaxial leads.

The $1 \times 10^{-8} R_H$ in-phase correction to R_H for 100 pF leads meets our desired $10^{-8} R_H$ absolute accuracy, but there is a 1 % contribution to $V_H(Y,Z)$ in the 90° out-of-phase j term. Auxiliary balances in the NIST high precision ac bridges are not capable of providing out-of-phase adjustment signals larger than $5 \times 10^{-4} R_H$, so the 1 % out-of-phase signal is unacceptable. The approximate solutions are listed in the next subsection to show the source of this out-of-phase problem.

7.3 Approximate Double-Series Solutions

Some of the terms in the following approximate solutions were obtained using the results of the dc double-series analysis of [22]. Most terms were found in a tedious process by changing the individual values of circuit element components by an order of magnitude in the computer program, observing the calculated results, and then finding the algebraic expressions that produced these results. The approximate solutions yield numerical results that agree with the exact numerical results listed in Eqs. (29) and (30) to within at least two significant figures for both the real and imaginary parts, but other terms may need to be added to these approximate equations if the circuit components have values significantly different from those listed in Eqs. (7) and (28).

$$I_{C_{Ot}} \approx I_{C_{Ot_a}} = \{[\omega^2 C_{Ot} C_{Ot} r_{Ot} r_{Ot}] + j[\omega C_{Ot} r_{Ot}]\} I_{Ot} \quad (32a)$$

$$I_{r_{Ot}} \approx I_{r_{Ot_a}} = \{1 - j[\omega C_{Ot} r_{Ot}]\} I_{Ot} \quad (32b)$$

$$I_S \approx I_{S_a} = \left\{ \left[1 - \frac{r_S}{R_H} \right] + j[\omega C_6(r_S + r_d) - \omega C_5 r_c - \omega C_{Ot} r_{Ot}] - j \left[\omega \frac{L_S}{R_H} - \omega \frac{L_4}{R_H} \frac{r_S}{R_H} \right] \right\} I_{Ot} \quad (32c)$$

$$I_{C_B} \approx I_{C_{B_a}} = \{-[\omega^2 C_B(C_6 r_S - C_{Ot} r_{Ot}) r_S + \omega^2 C_B L_S] + j[\omega C_B r_S]\} I_{Ot} \quad (32d)$$

$$I_{K_B} \approx I_{K_{B_a}} = \left\{ \left[\frac{r_S}{r_{K_B}} \right] + j \left[\omega \frac{L_S}{r_{K_B}} \right] \right\} I_{Ot} \quad (32e)$$

$$I_d \approx I_{d_a} = \left\{ \left[1 - \frac{r_S}{R_H} \right] + j \left[\omega(C_B + C_6) r_S - \omega \frac{L_S}{R_H} \right] \right\} I_{Ot} \quad (32f)$$

$$I_{C_6} \approx I_{C_{6_a}} = \{[\omega^2 C_6 C_6 R_H(r_S + r_d) + \omega^2 C_6 C_6 r_S r_6 - \omega^2 C_6 L_S] + j[\omega C_6 r_S]\} I_{Ot} \quad (32g)$$

$$I_{C_5} \approx I_{C_{5_a}} = \{[\omega^2 C_5 C_5 R_H r_S - \omega^2 C_5 C_6 R_H(r_S + r_d) - \omega^2 C_B C_5 R_H(r_S + r_d)] + [\omega^2 C_6 C_6 R_H r_c] + j[\omega C_5 R_H]\} I_{Ot} \quad (32h)$$

$$I_c \approx I_{c_a} = I_{d_a} + I_{C_{5_a}} + I_{C_{6_a}} \quad (32i)$$

$$I_4 \approx I_{4_a} = \{1 - j[\omega C_{Ot} r_{Ot}]\} I_{Ot} - I_{S_a} \quad (32j)$$

$$I_{3'} \approx I_{3'_a} = \left\{ \left[\frac{r_D}{R_H} + \frac{(r_b + r_d)}{R_H} + \omega^2 C_1(C_1 - C_5) R_H R_H + \omega^2 C_1 C_1 R_H(r_1 - r_3) \right] - [\omega^2 C_1 C_6 R_H r_S + \omega^2(C_A + C_5) L_D] \right. \\ \left. + j[\omega C_1(R_H - r_3) + \omega(C_A + C_5) r_D] \right\} I_{Ot} \quad (32k)$$

$$I_{C_3} \approx I_{C_{3_a}} = \{-[\omega^2 C_3 C_5 R_H R_H + \omega^2 C_1 C_3 R_H r_3 + \omega^2 C_3 C_6 R_H r_S] + j[\omega C_3 R_H]\} I_{Ot} \quad (32l)$$

$$I_3 \approx I_{3_a} = I_{3'_a} + I_{C_{3_a}} \quad (32m)$$

$$I_b \approx I_{b_a} = I_{c_a} + I_{4_a} - I_{3'_a} \quad (32n)$$

$$I_{C_2} \approx I_{C_{2_a}} = \{[\omega^2 C_2(C_2 + C_6 - C_1) r_S r_4 + \omega^2 C_1 C_2 R_H r_b] + \left[\omega^2 C_2(C_2 - C_5) R_H r_b - \omega^2 C_2 L_S \frac{r_4}{R_H} \right] + j \left[\omega C_2 r_b + \omega C_2 \frac{r_S r_4}{R_H} \right]\} I_{Ot} \quad (32o)$$

$$I_{C_1} \approx I_{C_{1a}} = \left\{ [\omega^2 C_1 (C_1 - C_5) R_H R_H + \omega^2 C_1 C_1 R_H (r_1 - r_3)] \right. \\ \left. + \left[-\omega^2 C_1 C_6 R_H r_5 + \omega^2 C_1 L_D + \omega^2 C_1 L_1 \frac{r_5}{R_H} \right] \right. \\ \left. - \left[\omega^2 (C_1 + C_5) L_S \frac{r_5}{R_H} \right] + j[\omega C_1 R_H] \right\} I_{Ot} \quad (32p)$$

$$I_a \approx I_{a_a} = I_{b_a} + I_{C_{1a}} + I_{C_{2a}} \quad (32q)$$

$$I_{K_A} \approx I_{K_{Aa}} = \left\{ \left[\frac{R_H}{r_{K_A}} \right] + j \left[\omega C_5 \frac{R_H R_H}{r_{K_A}} \right] \right\} I_{Ot} \quad (32r)$$

$$I_{C_A} \approx I_{C_{Aa}} = \{ -[\omega^2 C_A C_5 R_H R_H] + j[\omega C_A R_H] \} I_{Ot} \quad (32s)$$

$$I_D \approx I_{D_a} = I_{a_a} + I_{K_{Aa}} + I_{C_{Aa}} \quad (32t)$$

$$I_{C_D} \approx I_{C_{Da}} = \{ -[\omega^2 C_D C_5 R_H R_H] + j[\omega C_D R_H] \} I_{Ot} \quad (32u)$$

$$I_{C_{Pt}} \approx I_{C_{Pta}} = \{ -[\omega^2 C_{Pt} C_5 R_H R_H] + j[\omega C_{Pt} R_H] \} I_{Ot} \quad (32v)$$

$$I_{D_r} \approx I_{D_{ra}} = I_{D_a} + I_{3_a} + I_{C_{Da}} + I_{C_{Pta}} \quad (32w)$$

$$I_{C_{Dr}} \approx I_{C_{Dra}} = \{ -[\omega^2 C_{Dr} C_5 R_H R_H] + j[\omega C_{Dr} R_H] \} I_{Ot} \quad (32x)$$

$$I_{Dr} \approx I_{D_{ra}} = I_{D_{ra}} + I_{C_{Dra}} \quad (32y)$$

As expected, Eqs. (32i) and (32h) suggest that the current I_{C_5} in Fig. 3 enters the Drive, goes to point Y, to point D', through longitudinal resistances r_a , r_b , and r_c , through arm 5, and then exits through capacitance-to-shield C_5 . We would have likewise assumed that the current I_{C_1} enters the Drive, goes to point Y, to point D', through r_a , through arm 1, and then exits through C_1 . However, the approximate Eqs. (32p) and (32k), where I_{C_1} appears in $I_{3'}$, suggest that I_{C_1} enters the Drive, goes to point Y, to point 3, through arm 3, travels "upstream" through r_b , through arm 1, and then exits through C_1 . The current I_{C_3} , on the other hand, enters the Drive, goes to point Y, to point 3, and then exits through C_3 , bypassing the device altogether; this latter effect provides an advantage to double-series connections by reducing shunt currents within the device.

Expressing Eq. (27a) in terms of Eq. (27b), we find that Eq. (33) gives the approximate quantum Hall voltage correction terms.

We see from Eq. (33) that sample probe lead 5, just as in the single-series "normal" case, is the dominant source of the 1 % out-of-phase component of the quantum Hall voltage signal in the numerical example for this double-series connection to the QHE device. The next subsection investigates the effect of removing this lead, which was effective before in the single-series "normal" case of Sec. 5.

$$\Delta_{YZ} \approx \left\{ \left[\frac{r_5 r_4}{R_H R_H} + \frac{r_3 r_D}{R_H R_H} + \frac{r_3 (r_b + r_a)}{R_H R_H} + \omega^2 C_5 C_5 R_H r_5 \right] \right. \\ - [\omega^2 C_B C_5 R_H r_5 + \omega^2 C_6 (C_5 - C_6) R_H r_5] \\ - \omega^2 C_1 (C_1 - C_5) R_H r_3] \\ + [\omega^2 C_6 C_6 R_H (r_d + r_c) - \omega^2 C_5 (C_B + C_6) R_H r_d] \\ + [\omega^2 C_6 C_6 r_5 r_6 - \omega^2 C_1 C_6 r_5 r_3 + \omega^2 C_1 C_1 r_3 (r_1 - r_3)] \\ - \left[\omega^2 C_B L_S + \omega^2 C_6 (L_S + L_3) + \omega^2 (C_A + C_5) L_D \frac{r_3}{R_H} \right] \\ \left. + j[\omega C_5 (R_H + r_d) + \omega C_1 r_3 + \omega (C_B + C_6) r_5] \right. \\ \left. - j \left[\omega C_6 (r_5 + r_d) + \omega L_4 \frac{r_5}{R_H R_H} \right] \right\}. \quad (33)$$

7.4 Disconnecting Sample Probe Lead 5

Equation (33) predicts that the out-of-phase term $j[\omega C_5 R_H]$ in the expression for Δ_{YZ} can be reduced by disconnecting coaxial cable 5 at position 5', where 5' is either located at the potential contact pad on the QHE device, or at an intermediate contact point in the sample holder. There is a capacitance $C_{5'}$ between the QHE device and the shield that replaces capacitance C_5 in Fig. 3. Also, a shield impedance $z_{5'}$ replaces the lead impedance z_5 .

If we assume in the numerical examples that

$$C_{5'} = C_A = C_B = 1 \text{ pF} \quad (34a)$$

$$r_{5'} = r_5 = 1 \text{ } \Omega. \quad (34b)$$

Then

$$\Delta_{YZ} = \{ [2.0 \times 10^{-8}] + j[1.0 \times 10^{-4}] \} \quad (35)$$

when the coaxial lead capacitances are all 100 pF, and

$$\Delta_{YZ} = \{ [9.4 \times 10^{-8}] + j[1.1 \times 10^{-4}] \} \quad (36)$$

when they are 250 pF. All experiments which have measured ac values of $V_H(Y,Z)$ for double-series connections have had to remove coaxial lead 5 because of the effects due to the large capacitance-to-shield C_5 presented above.

Equation (33) could be used to apply corrections to the experimental data in order to reduce the 9.4×10^{-8} in-phase error in $R_H I_{Ot}$. However, our approximate and exact equations do not include the effects of wire-to-wire capacitances; the bridge auxiliary balance could

introduce unintentional in-phase contributions because of the large out-of-phase component of $V_H(Y,Z)$; and it is not trivial to measure the value of C_5 in order to apply the correction with lead 5 disconnected.

7.5 Double-Series Connections at the QHE Device

Many experiments have made double-series connections to the QHE device at the bottom of the sample probe by using short bonding wires to form the circuit. Points Y and Z of Fig. 3 are thus moved from outside the sample probe down to the sample holder. There are no coaxial leads connected to points 1, 2, 5, and 6, so their capacitances-to-shield become much smaller. Four coaxial leads labeled Ot, Dt, Pt, and Dr connect the QHE device to the outside world. The double-series circuit shown in Fig. 3 remains exactly the same for this case, as do Eqs. (24) through (27). The values of some circuit components, however, change.

We use the following cardinal values in our numerical example

$$R_H = 10^4 \Omega \quad (37a)$$

$$r_S = r_1 = r_2 = r_3 = r_4 = r_5 = r_6 = r_D = 10^{-3} \Omega \quad (37b)$$

$$r_a = r_b = r_c = r_d = 10^{-3} \Omega \quad (37c)$$

$$r_{K_A} = r_{K_B} = 10^{12} \Omega \quad (37d)$$

$$r_{Ot} = r_{Dt} = r_{Pt} = r_{Dr} = 1 \Omega \quad (37e)$$

$$C_S = C_1 = C_2 = C_3 = C_4 = C_5 = C_6 = C_D = 10^{-12} \text{ F} \quad (37f)$$

$$C_A = C_B = 10^{-12} \text{ F} \quad (37g)$$

$$C_{Ot} = C_{Dt} = C_{Pt} = C_{Dr} = 10^{-10} \text{ F} \quad (37h)$$

The numerical results for the currents in Eqs. (26) with the double-series leads connected at the bottom of the sample probe are

$$I_{C_{Ot}}^B = \{[1.0 \times 10^{-12}] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (38a)$$

$$I_{r_{Ot}}^B = \{[1.000000] - j[1.0 \times 10^{-6}]\} I_{Ot} \quad (38b)$$

$$I_S^B = \{[1.000000] - j[2.0 \times 10^{-6}]\} I_{Ot} \quad (38c)$$

$$I_{C_B}^B = \{-[1.0 \times 10^{-10}] + j[1.0 \times 10^{-11}]\} I_{Ot} \quad (38d)$$

$$I_{K_B}^B = \{[1.0 \times 10^{-15}] + j[1.0 \times 10^{-14}]\} I_{Ot} \quad (38e)$$

$$I_d^B = \{[1.000000] - j[2.0 \times 10^{-6}]\} I_{Ot} \quad (38f)$$

$$I_{C_6}^B = \{-[1.0 \times 10^{-10}] + j[2.0 \times 10^{-11}]\} I_{Ot} \quad (38g)$$

$$I_{C_5}^B = \{[1.0 \times 10^{-10}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (38h)$$

$$I_c^B = \{[1.000000] + j[9.8 \times 10^{-5}]\} I_{Ot} \quad (38i)$$

$$I_4^B = \{[3.0 \times 10^{-7}] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (38j)$$

$$I_{3'}^B = \{[3.0 \times 10^{-7}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (38k)$$

$$I_{C_3}^B = \{-[1.0 \times 10^{-8}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (38l)$$

$$I_3^B = \{[2.9 \times 10^{-7}] + j[2.0 \times 10^{-4}]\} I_{Ot} \quad (38m)$$

$$I_b^B = \{[1.000000] - j[2.0 \times 10^{-6}]\} I_{Ot} \quad (38n)$$

$$I_{C_2}^B = \{[9.8 \times 10^{-16}] + j[1.0 \times 10^{-11}]\} I_{Ot} \quad (38o)$$

$$I_{C_1}^B = \{[2.0 \times 10^{-10}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (38p)$$

$$I_a^B = \{[1.000000] + j[9.8 \times 10^{-5}]\} I_{Ot} \quad (38q)$$

$$I_{K_A}^B = \{[1.0 \times 10^{-8}] + j[9.8 \times 10^{-13}]\} I_{Ot} \quad (38r)$$

$$I_{C_A}^B = \{-[9.8 \times 10^{-9}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (38s)$$

$$I_D^B = \{[1.000000] + j[2.0 \times 10^{-4}]\} I_{Ot} \quad (38t)$$

$$I_{C_D}^B = \{-[9.9 \times 10^{-9}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (38u)$$

$$I_{C_{Pt}}^B = \{-[9.8 \times 10^{-7}] + j[0.010000]\} I_{Ot} \quad (38v)$$

$$I_{r_{Dr}}^B = \{[1.000000] + j[0.010500]\} I_{Ot} \quad (38w)$$

$$I_{C_{Dr}}^B = \{-[1.0 \times 10^{-6}] + j[0.010000]\} I_{Ot} \quad (38x)$$

$$I_{Dr}^B = \{[1.000000] + j[0.020500]\} I_{Ot} \quad (38y)$$

Expressing Eq. (27a) in the form of Eq. (27b), we find that

$$V_H^B(Y,Z) = \{1 - [1.0 \times 10^{-10}] + j[9.8 \times 10^{-5}]\} R_H I_{Ot}, \quad (39a)$$

$$\Delta_{YZ}^B = \{-[1.0 \times 10^{-10}] + j[9.8 \times 10^{-5}]\} \quad (39b)$$

for 100 pF lead capacitances and

$$\Delta_{YZ}^B = \{[1.8 \times 10^{-10}] + j[9.5 \times 10^{-5}]\} \quad (40)$$

for 250 pF coaxial leads.

Equation (40) implies a very small in-phase error in $R_H I_{Ot}$. This is not supported by measurements, which have observed errors in $R_H I_{Ot}$ of order 10^{-7} . The discrepancy could either be due to unintentional in-phase

contributions from the bridge auxiliary balances arising from the large out-of-phase component of $V_H(Y,Z)$, or because our equations do not include the effects of wire-to-wire capacitances.

To assist laboratories who are making double-series connection measurements at the QHE device we list the additional terms that should be added to the approximate current and quantum Hall voltage solutions given by Eqs. (32) and (33). Those equations which require additional terms are:

$$I_d^B \approx I_{d_a}^B = I_{d_a} - \{j[\omega C_{Ot} r_{Ot}]\} I_{Ot} \quad (41a)$$

$$I_{C_6}^B \approx I_{C_{6a}}^B = I_{C_{6a}} + \{j[\omega C_6 r_d]\} I_{Ot} \quad (41b)$$

$$I_{C_5}^B \approx I_{C_{5a}}^B = I_{C_{5a}} + \{[\omega^2 C_{Ot} C_5 R_H r_{Ot}]\} I_{Ot} \quad (41c)$$

$$I_4^B \approx I_{4a}^B = I_{4a} + \left\{ \left[\frac{(r_d + r_c)}{R_H} \right] \right\} I_{Ot} \quad (41d)$$

$$I_{3'}^B \approx I_{3'a}^B = I_{3'a} + \left\{ j \left[\omega \frac{L_D}{R_H} \right] \right\} I_{Ot} \quad (41e)$$

$$I_{K_A}^B \approx I_{K_{Aa}}^B = I_{K_{Aa}} - \left\{ j \left[\omega C_{Ot} \frac{R_H}{r_{K_A}} r_{Ot} + \omega \frac{L_D}{r_{K_A}} \right] \right\} I_{Ot} \quad (41f)$$

$$I_{C_A}^B \approx I_{C_{Aa}}^B = I_{C_{Aa}} + \{[\omega^2 C_{Ot} C_A R_H r_{Ot} + \omega^2 C_A L_D]\} I_{Ot} \quad (41g)$$

$$I_{C_D}^B \approx I_{C_{Da}}^B = I_{C_{Da}} + \{[\omega^2 C_{Ot} C_D R_H r_{Ot}]\} I_{Ot} \quad (41h)$$

$$I_{C_{Pt}}^B \approx I_{C_{Pta}}^B = I_{C_{Pta}} + \{[\omega^2 C_{Ot} C_{Pt} R_H r_{Ot} + \omega^2 C_{Pt} C_{Pt} R_H r_{Pt}]\} I_{Ot} \quad (41i)$$

and

$$\begin{aligned} \Delta_H^B \approx & \Delta_H + \{[\omega^2 C_5 C_{Ot} R_H r_{Ot} + \omega^2 C_5 C_{Pt} R_H r_{Pt}] \\ & - \omega^2 C_{Ot} C_{Ot} r_{Ot} r_{Ot}\} - \{[\omega^2 C_{Pt} C_{Pt} r_{Pt} r_{Pt}] \\ & - j[\omega C_{Ot} r_{Ot} + \omega C_{Pt} r_{Pt}]\}. \end{aligned} \quad (42)$$

We once again caution the reader that these approximate equations do not include the effects of wire-to-wire capacitances. This circuit is not a good candidate for further analysis because the quantized Hall and longitudinal resistances could not be measured on the same cool-down.

8. Triple-Series Circuit

The double-series circuit of Fig. 3 could be converted to a triple-series circuit by adding short coaxial leads

between points Y and 1 and Z and 6. We do not consider this triple-series circuit since it would involve several additional months of effort to perform the analysis, and the problems found in double-series circuits in Sec. 7 due to large shunt currents through C_5 also occur in triple-series circuits. Either coaxial lead 5 would have to be disconnected at position 5' at the QHE device end of the sample probe, or the triple-series connections would have to be made at the device. Neither choice satisfies our goal of measuring the ac and dc quantized Hall and longitudinal resistances on the same cool-down. We therefore proceed to quadruple-series connections, which turns out to satisfy our requirements at this stage of analysis.

9. Analysis of the Quadruple-Series Circuit

Figure 4 shows an equivalent electrical circuit representation of an ac QHRS using two quadruple-series connections to the QHE device. It is quadruple-series because short coaxial leads outside the sample probe connect room temperature access points 5, 3, 1, and D at point Y, providing four current paths to the device. Other short coaxial leads connect access points 2, 4, 6, and S at point Z. Short coaxial leads outside the sample probe connect point Y with the Drive and Potential ports, and point Z with the Inner/Outer and Detector ports.

9.1 Exact Quadruple-Series Equations

To simplify the final algebraic expressions, we once again define substitutions of variables, and substitutions of substitutions. Let

$$A = j\omega C_{Ot} r_{Ot} \quad (43a)$$

$$B = j\omega C_B z_S \quad (43b)$$

$$C = \frac{1}{j\omega C_B r_{KB}} \quad (43c)$$

$$D_1 = \frac{z_S}{(R_H + z_6)} \quad (43d)$$

$$D_2 = \frac{r_d}{(R_H + z_6)} \quad (43e)$$

$$D_3 = \frac{(R_H + r_d)}{z_5} \quad (43f)$$

$$D_4 = \frac{1}{j\omega C_B z_5} \quad (43g)$$




Fig. 4. An equivalent electrical circuit representation of an ac QHE resistance standard with two quadruple-series connections to the device. See Sec. 9 for the circuit analysis.

$$D_5 = \frac{1}{j\omega C_5 z_5} \quad (43h)$$

$$D_6 = \frac{z_6}{(R_H + z_4)} \quad (43i)$$

$$D_7 = \frac{r_c}{(R_H + z_4)} \quad (43j)$$

$$D_8 = \frac{C_3}{C_5} \quad (43k)$$

$$E_1 = \frac{z_4}{(R_H + z_2)} \quad (43l)$$

$$E_2 = \frac{r_b}{(R_H + z_2)} \quad (43m)$$

$$E_3 = \frac{C_1}{C_3} \quad (43n)$$

$$E_4 = \frac{(R_H + r_a)}{r_{K_A}} \quad (43o)$$

$$E_5 = \frac{z_2}{r_{K_A}} \quad (43p)$$

$$E_6 = j\omega C_A r_{K_A} \quad (43q)$$

$$F_1 = \frac{(R_H + z_5)}{z_3} \quad (43r)$$

$$F_2 = \frac{r_c}{z_3} \quad (43s)$$

$$F_3 = \frac{(R_H + z_3)}{z_1} \quad (43t)$$

$$F_4 = \frac{r_b}{z_1} \quad (43u)$$

$$F_5 = \frac{(R_H + z_1)}{z_D} \quad (43v)$$

$$F_6 = \frac{r_a}{z_D} \quad (43w)$$

$$F_7 = \frac{C_5}{C_A} \quad (43x)$$

$$F_8 = j\omega C_5 z_D \quad (43y)$$

$$G_1 = F_8 + E_6(F_7 + F_8) \quad (43z)$$

$$G_2 = F_8 + G_1(E_4 + E_5) \quad (44a)$$

$$G_3 = F_8 + E_4 G_1 \quad (44b)$$

$$G_4 = G_3(1 + F_4) + E_2 G_2 \quad (44c)$$

$$G_5 = G_4 + F_3 G_3 \quad (44d)$$

$$G_6 = G_4 + E_1 G_2 \quad (44e)$$

$$G_7 = G_4 + F_2 G_5 + D_7 G_6 \quad (44f)$$

$$G_8 = G_7 + D_6 G_6 \quad (44g)$$

$$G_9 = G_7 + F_1 G_5 \quad (44h)$$

$$H_1 = \frac{(D_5 G_7 - D_3)}{(1 + D_5 G_9)} \quad (44i)$$

$$H_2 = \frac{D_4}{(1 + D_5 G_9)} \quad (44j)$$

$$H_3 = \frac{D_5 G_8}{(1 + D_5 G_9)} \quad (44k)$$

$$J_1 = 1 + E_1 + E_2 \quad (44l)$$

$$J_2 = E_2(1 + F_2) \quad (44m)$$

$$J_3 = J_2 + D_7 J_1 \quad (44n)$$

$$J_4 = 1 + J_3 + D_6 J_1 \quad (44o)$$

$$J_5 = J_3 + E_2 F_1 \quad (44p)$$

$$J_6 = J_4 - H_3 J_5 \quad (44q)$$

$$J_7 = J_3 - H_1 J_5 + D_2 J_6 \quad (44r)$$

$$J_8 = 1 + J_7 + D_1 J_6 \quad (44s)$$

$$J_9 = J_7(1 + C) + H_2 J_5 \quad (44t)$$

$$J_{10} = \frac{1}{(J_8 + B J_9)} \quad (44u)$$

$$K_1 = \frac{C_D}{C_A} \quad (44v)$$

$$K_2 = j\omega C_D z_D \quad (44w)$$

$$K_3 = \frac{C_{Pt}}{C_D} \frac{1}{[1 + j\omega C_{Pt} r_{Pt}]} \quad (44x)$$

$$K_4 = \frac{C_{Dr}}{C_D} \quad (44y)$$

$$K_5 = j\omega C_{Dr} r_{Dr}. \quad (44z)$$

Eight of the current solutions are trivial because of the four-terminal-pair definition [15,16]

$$I_{Dt} = I_{Pt} = I_{Cs} = I_{C_{Dt}} = I_{C_6} = I_{C_4} = I_{C_2} = I_{r_{Dt}} = 0. \quad (45a)$$

The remaining exact equations for the quadruple-series circuit currents are

$$I_{C_{Ot}} = \frac{A}{(1+A)} I_{Ot} \quad (45b)$$

$$I_{r_{Ot}} = I_{Ot} - I_{C_{Ot}} \quad (45c)$$

$$I_S = J_{10} I_{r_{Ot}} \quad (45d)$$

$$I_{C_B} = B I_S \quad (45e)$$

$$I_{K_B} = C I_{C_B} \quad (45f)$$

$$I_d = I_S + I_{C_B} + I_{K_B} \quad (45g)$$

$$I_6 = D_1 I_S + D_2 I_d \quad (45h)$$

$$I_{5'} = H_1 I_d - H_2 I_{C_B} + H_3 I_6 \quad (45i)$$

$$I_{C_5} = G_7 I_d + G_8 I_6 - G_9 I_{5'} \quad (45j)$$

$$I_5 = I_{5'} + I_{C_5} \quad (45k)$$

$$I_c = I_d - I_{5'} + I_6 \quad (45l)$$

$$I_4 = D_6 I_6 + D_7 I_c \quad (45m)$$

$$I_{3'} = F_1 I_{5'} - F_2 I_c \quad (45n)$$

$$I_{C_3} = D_8 I_{C_5} \quad (45o)$$

$$I_3 = I_{3'} + I_{C_3} \quad (45p)$$

$$I_b = I_c - I_{3'} + I_4 \quad (45q)$$

$$I_2 = E_1 I_4 + E_2 I_b \quad (45r)$$

$$I_{1'} = F_3 I_{3'} - F_4 I_b \quad (45s)$$

$$I_{C_1} = E_3 I_{C_3} \quad (45t)$$

$$I_1 = I_{1'} + I_{C_1} \quad (45u)$$

$$I_a = I_b - I_{1'} + I_2 \quad (45v)$$

$$I_{K_A} = E_4 I_a + E_5 I_2 \quad (45w)$$

$$I_{C_A} = E_6 I_{K_A} \quad (45x)$$

$$I_D = I_a + I_{K_A} + I_{C_A} \quad (45y)$$

$$I_{C_D} = K_1 I_{C_A} + K_2 I_D \quad (45z)$$

$$I_{C_{Pt}} = K_3 I_{C_D} \quad (46a)$$

$$I_{r_{Dr}} = I_D + I_1 + I_3 + I_5 + I_{C_D} + I_{C_{Pt}} \quad (46b)$$

$$I_{C_{Dr}} = K_4 I_{C_D} + K_5 I_{r_{Dr}} \quad (46c)$$

$$I_{Dr} = I_{r_{Dr}} + I_{C_{Dr}} \quad (46d)$$

The exact equation for the quantum Hall voltage is obtained by summing the voltages between the inner conductors of the Detector coaxial port and the Potential coaxial port. Taking the path through point Z, arm 4, voltage generators V_{c4} and V_{c3} , arm 3, and point Y we find that

$$V_H(Y,Z) = R_H I_c + (R_H + z_4) I_4 + z_3 I_{3'} - r_{Pt} I_{C_{Pt}}, \quad (47a)$$

which can also be expressed as

$$V_H(Y,Z) = [1 + \Delta_{YZ}] R_H I_{Ot}. \quad (47b)$$

9.2 A Numerical Example

We investigate the parasitic impedance contributions of the ac QHRS on the measured value of $V_H(Y,Z)$ for a particular example of the quadruple-series circuit by using the cardinal numbers listed in Eqs. (7) and Eqs. (28). The numerical results for the currents in Eqs. (45) and (46) are

$$I_{C_{Ot}} = \{[1.0 \times 10^{-22}] + j[1.0 \times 10^{-11}]\} I_{Ot} \quad (48a)$$

$$I_{r_{Ot}} = \{[1.00000] - j[1.0 \times 10^{-11}]\} I_{Ot} \quad (48b)$$

$$I_S = \{[0.99990] - j[1.0 \times 10^{-6}]\} I_{Ot} \quad (48c)$$

$$I_{C_B} = \{-[1.0 \times 10^{-10}] + j[1.0 \times 10^{-8}]\} I_{Ot} \quad (48d)$$

$$I_{K_B} = \{[1.0 \times 10^{-12}] + j[1.0 \times 10^{-14}]\} I_{Ot} \quad (48e)$$

$$I_d = \{[0.99990] - j[9.9 \times 10^{-7}]\} I_{Ot} \quad (48f)$$

$$I_6 = \{[1.0 \times 10^{-4}] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (48g)$$

$$I_{5'} = \{[1.0 \times 10^{-7}] + j[3.1 \times 10^{-14}]\} I_{Ot} \quad (48h)$$

$$I_{C_5} = \{-[1.0 \times 10^{-10}] + j[0.01000]\} I_{Ot} \quad (48i)$$

$$I_5 = \{[1.0 \times 10^{-7}] + j[0.01000]\} I_{Ot} \quad (48j)$$

$$I_c = \{[1.00000] + j[9.8 \times 10^{-9}]\} I_{Ot} \quad (48k)$$

$$I_4 = \{[1.1 \times 10^{-7}] + j[2.0 \times 10^{-10}]\} I_{Ot} \quad (48l)$$

$$I_{3'} = \{[1.1 \times 10^{-7}] + j[2.0 \times 10^{-10}]\} I_{Ot} \quad (48m)$$

$$I_{C_3} = \{-[1.0 \times 10^{-10}] + j[0.01000]\} I_{Ot} \quad (48n)$$

$$I_3 = \{[1.1 \times 10^{-7}] + j[0.01000]\} I_{Ot} \quad (48o)$$

$$I_b = \{[1.00000] + j[9.8 \times 10^{-9}]\} I_{Ot} \quad (48p)$$

$$I_2 = \{[1.0 \times 10^{-7}] + j[3.1 \times 10^{-14}]\} I_{Ot} \quad (48q)$$

$$I_{1'} = \{[1.0 \times 10^{-4}] + j[1.0 \times 10^{-6}]\} I_{Ot} \quad (48r)$$

$$I_{C_1} = \{-[1.0 \times 10^{-10}] + j[0.01000]\} I_{Ot} \quad (48s)$$

$$I_1 = \{[1.0 \times 10^{-4}] + j[0.01000]\} I_{Ot} \quad (48t)$$

$$I_a = \{[0.99990] - j[1.0 \times 10^{-6}]\} I_{Ot} \quad (48u)$$

$$I_{K_A} = \{[1.0 \times 10^{-8}] - j[1.0 \times 10^{-14}]\} I_{Ot} \quad (48v)$$

$$I_{C_A} = \{[1.0 \times 10^{-10}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (48w)$$

$$I_D = \{[0.99990] + j[9.9 \times 10^{-5}]\} I_{Ot} \quad (48x)$$

$$I_{C_D} = \{-[1.0 \times 10^{-10}] + j[0.01000]\} I_{Ot} \quad (48y)$$

$$I_{C_{Pt}} = \{-[1.0 \times 10^{-12}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (48z)$$

$$I_{Dr} = \{[1.00000] + j[0.04020]\} I_{Ot} \quad (49a)$$

$$I_{C_{Dr}} = \{-[1.4 \times 10^{-12}] + j[1.0 \times 10^{-4}]\} I_{Ot} \quad (49b)$$

$$I_{Dr} = \{[1.00000] + j[0.04030]\} I_{Ot} \quad (49c)$$

The 90° out-of-phase parts of shunt currents I_{C_5} , I_{C_3} , I_{C_1} , I_{C_D} , I_{C_A} , $I_{C_{Pt}}$, and $I_{C_{Dr}}$ are much larger than for shunt currents I_{C_B} and $I_{C_{Ot}}$ because contact pads 5', 3', 1', and D' are all near the quantum Hall potential, rather than near the shield potential. A 1 % out-of-phase current passes through each of the coaxial cable capacitances C_5 , C_3 , C_1 , and C_D in this numerical example, which is not necessarily a problem if the bridge Drive can provide this extra 4 % of out-of-phase current to I_{Dr} . We can see from the small out-of-phase components of currents $I_{5'}$, $I_{3'}$, $I_{1'}$, and I_D that the four shunt currents I_{C_5} , I_{C_3} , I_{C_1} , and I_{C_D} all bypass the QHE device, which is a great advantage of the quadruple-series circuit.

Expressing Eq. (47a) in the form of Eq. (47b), we find that

$$V_H(Y,Z) = \{1 - [2.0 \times 10^{-7}] + j[1.0 \times 10^{-8}]\} R_H I_{Ot}, \quad (50a)$$

$$\Delta_{YZ} = \{-[2.0 \times 10^{-7}] + j[1.0 \times 10^{-8}]\} \quad (50b)$$

for 100 pF lead capacitances, and also the same value

$$\Delta_{YZ} = \{-[2.0 \times 10^{-7}] + j[1.0 \times 10^{-8}]\} \quad (51)$$

for 250 pF coaxial leads.

There is only a $1 \times 10^{-8} R_H I_{Ot}$ out-of-phase component in the $V_H(Y,Z)$ signal for the numerical examples given in Eqs. (50) and (51). Unlike the double-series circuit, this out-of-phase result is very promising. The real part of $V_H(Y,Z)$, however, appears to have a very large error term that is $-2 \times 10^{-7} R_H I_{Ot}$ in these two examples; but $V_H(Y,Z)$ actually is the quantized Hall voltage V_H across the device *minus* the longitudinal voltage $V_x(2,6)$ along the device between points 2 and 6,

$$V_H(Y,Z) \equiv V_H - V_x(2,6), \quad (52a)$$

and

$$V_x(2,6) \approx (r_b + r_c) I_{Ot}. \quad (52b)$$

Therefore,

$$V_H \approx \left[1 + \Delta_H + \frac{(r_b + r_c)}{R_H} \right] R_H I_{Ot}, \quad (53a)$$

or

$$V_H \approx [1 + \delta_H] R_H I_{Ot}. \quad (53b)$$

V_H has a correction factor δ_H in the real term that is only $-7.9 \times 10^{-11} R_H I_{Ot}$ for these two numerical examples where $(r_b + r_c) = 2 \times 10^{-7} R_H$.

The quadruple-series circuit is an excellent candidate as a possible ac QHRS, and will be further considered in a future paper which includes the effects of wire-to-wire capacitances. The approximate equations for the currents and quantum Hall voltage will be given in that paper. In the meantime, some approximate equations can found in Ref. [23].

10. Conclusions

We have used an equivalent electrical circuit model of the quantum Hall effect device to calculate the effects of parasitic impedances that are present in four-terminal-

pair [15,16] measurements of ac quantized Hall resistance standards. The discrete circuit components include all of the parasitic capacitances, inductances, and leakage resistances of the standard except the wire-to-wire-capacitances.

Exact algebraic equations have been derived for the currents and quantum Hall voltages for single-series “normal”, single-series “offset”, double-series, and quadruple-series circuit connections to the device. We find that the single-series “offset” and quadruple-series connections appear to meet our desired goals of measuring both the quantized Hall resistance R_H and the longitudinal resistance R_x in the same cool-down for both ac and dc currents with an absolute accuracy of $10^{-8} R_H$ or better. These two circuits will be further considered in a future paper in which the effects of wire-to-wire capacitances are also included in the analysis.

Acknowledgments

We thank Mr. John Matthews of the Department of Physics, University of Maryland, Drs. Kevin C. Lee, Michael H. Kelley, and Martin Misakian of the Electricity Division of NIST, and Dr. Jin S. Kim of the Semiconductor Electronics Division of NIST for their discussions and comments.

11. References

- [1] K. von Klitzing, G. Dorda, and M. Pepper, New Method for High-Accuracy Determination of the Fine-Structure Constant Based on Quantized Hall Resistance, *Phys. Rev. Lett.* **45**, 494–497 (1980).
- [2] The Quantum Hall Effect, R. E. Prange and S. M. Girvin, eds., Springer-Verlag, New York (1987) pp. 1–419.
- [3] The Integral and Fractional Quantum Hall Effects, C. T. Van Degrift, M. E. Cage, and S. M. Girvin, eds., American Association of Physics Teachers, College Park, Maryland (1991) pp. 1–116.
- [4] F. Delahaye, Accurate AC Measurements of the Quantized Hall Resistance from 1 Hz to 1.6 kHz, *Metrologia* **31**, 367–373 (1995).
- [5] B. M. Wood, A. D. Inglis, and M. Cote, Evaluation of the AC Quantized Hall Resistance, *IEEE Trans. Instrum. Meas.* **46**, 269–272 (1997).
- [6] J. Melcher, P. Warnecke, and R. Hanke, Comparison of Precision AC and DC Measurements with the Quantized Hall Resistance, *IEEE Trans. Instrum. Meas.* **42**, 292–294 (1993).
- [7] A. Hartland, B. P. Kibble, P. J. Rodgers, and J. Bohacek, AC Measurements of the Quantized Hall Resistance, *IEEE Trans. Instrum. Meas.* **44**, 245–248 (1995).
- [8] J. Bohacek, P. Svoboda, and P. Vasek, AC QHE-Based Calibration of Resistance Standards, *IEEE Trans. Instrum. Meas.* **46**, 273–275 (1997).
- [9] A. Hartland, B. P. Kibble, and S. W. Chua, AC Quantized Hall Resistance Experiments at the National Physical Laboratory, BEMC '97 Conference Digest, National Physical Laboratory, Teddington, U. K. (1997) pp. 18-1 to 18-4.
- [10] F. Delahaye, Technical Guidelines for Reliable Measurements of the Quantized Hall Resistance, *Metrologia* **26**, 63–68 (1989).
- [11] A. Jeffery, R. E. Elmquist, J. Q. Shields, L. H. Lee, M. E. Cage, S. H. Shields, and R. F. Dziuba, Determination of the von Klitzing Constant and the Fine-Structure Constant through a Comparison of the Quantized Hall Resistance and the Ohm Derived from the NIST Calculable Capacitor, *Metrologia* **35**, 83–96 (1998).
- [12] B. N. Taylor and T. J. Witt, New International Electrical Reference Standards Based on the Josephson and Quantum Hall Effects, *Metrologia* **26**, 47–62 (1989).
- [13] M. E. Cage, B. F. Field, R. F. Dziuba, S. M. Girvin, A. C. Gossard, and D. C. Tsui, Temperature Dependence of the Quantum Hall Resistance, *Phys. Rev. B* **30**, 2286–2288 (1984).
- [14] M. E. Cage, R. F. Dziuba, B. F. Field, E. R. Williams, S. M. Girvin, A. C. Gossard, D. C. Tsui, and R. J. Wagner, Dissipation and Dynamic Nonlinear Behavior in the Quantum Hall Regime, *Phys. Rev. Lett.* **51**, 1374–1377 (1983).
- [15] R. D. Cutkosky, Four-Terminal-Pair Networks as Precision Admittance and Impedance Standards, *Commun. Electron.* **70**, 19–22 (1964).
- [16] R. D. Cutkosky, Techniques for Comparing Four-Terminal-Pair Admittance Standards, *J. Res. Natl. Bur. Stand. (U.S.)* **74C**, 63–78 (1970).
- [17] D. N. Homan, Applications of Coaxial Chokes to AC Bridge Circuits, *J. Res. Natl. Bur. Stand. (U.S.)* **72C**, 161–165 (1968).
- [18] B. W. Ricketts and P. C. Kemeny, Quantum Hall Effect Devices as Circuit Elements, *J. Phys. D: Appl. Phys.* **21**, 483–487 (1988).
- [19] M. E. Cage and A. Jeffery, Intrinsic Capacitances and Inductances of Quantum Hall Effect Devices, *J. Res. Natl. Inst. Stand. Technol.* **101**, 733–744 (1996).
- [20] F. Piquemal, G. Trapon, and G. Geneves, AC Measurements of the Minimum Longitudinal Resistance of a QHE Sample from 10 Hz to 10 kHz, *IEEE Trans. Instrum. Meas.* **45**, 918–922 (1996).
- [21] F. Cabiati, L. Callegaro, C. Cassiago, V. D'Elia, and G. M. Reedtz, Effect of the Frequency on the Longitudinal Resistance of a GaAs-AlGaAs Heterostructure, *CCE/97-32*, 1–5 (1997).
- [22] M. E. Cage, A. Jeffery, R. E. Elmquist, and K. C. Lee, Calculating the Effects of Longitudinal Resistance in Multi-Series-Connected Quantum Hall Effect Devices, *J. Res. Natl. Inst. Stand. Technol.* **103**, 561–592 (1998).
- [23] M. E. Cage and A. Jeffery, A Problem in AC Quantized Hall Resistance Measurements and a Proposed Solution, *J. Res. Natl. Inst. Stand. Technol.* **103**, 593–604 (1998).
- [24] F. Delahaye, Series and Parallel Connection of Multiterminal Quantum Hall Effect Devices, *J. Appl. Phys.* **73**, 7915–7920 (1993).
- [25] A. Jeffery, R. E. Elmquist, and M. E. Cage, Precision Tests of a Quantum Hall Effect Device DC Equivalent Circuit using Double-Series and Triple-Series Connections, *J. Res. Natl. Inst. Stand. Technol.* **100**, 677–685 (1995).

About the authors: M. E. Cage and A. Jeffery are physicists in the Electricity Division of the NIST Electronics and Electrical Engineering Laboratory. The National Institute of Standards and Technology is an agency of the Technology Administration, U.S. Department of Commerce.